



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111754

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/29/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	6.99	10/29/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	10/29/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	10/29/2020	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	10/29/2020	16:16
6	L4222-11	1	Solid	20.03	20	19.4	8.07	10/29/2020	16:25
7	L4222-11DUP	1	Solid	20.04	20	19.5	8.08	10/29/2020	16:26
8	L4548-02	1	Solid	20.02	20	19.1	4.93	10/29/2020	16:29
9	L4568-01	1	Solid	20.01	20	19.6	7.71	10/29/2020	16:30
10	L4568-02	1	Solid	20.03	20	19.2	6.57	10/29/2020	16:31
11	CCV2	1	Water	NA	NA	20.1	12.02	10/29/2020	16:33

WORKLIST(Hardcopy Internal Chain)

WB 111754

WorkList Name : CORROSIONITY L4222

WorkList ID : 143755

Department : Wet-Chemistry

Date : 10-29-2020 14:26:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Solid	L4222-11	Corrosivity	Cool 4 deg C	PSEG05		EO-02-102920	10/29/2020	9045D
10/28/2020	Solid	L4568-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-WASH	10/28/2020	9045D
10/28/2020	Solid	L4568-02	Corrosivity	Cool 4 deg C	CLEA11		RECLAMATION-PIT	10/28/2020	9045D

Date/Time 10/29/2020 16:00
 Received by: Jn (wet Chem)
 Relinquished by: Cip (SM)

Date/Time 10/29/2020 17:35
 Received by: Cip (SM)
 Relinquished by: Jn (wet Chem)