

LB111758


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55980



A= 0.0000e+000

B= 1.7728e-004

C= 2.0820e-002

Rho= 0.9996736

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	47	0.000	47				
0.2	0.200	0.224	0.024	1144	0.0 %	1144				
2.5	2.500	2.475	-0.025	13846	0.0 %	13846				
5.0	5.000	4.859	-0.141	27291	0.0 %	27291				
7.5	7.500	7.668	0.168	43139	0.0 %	43139				
10.0	10.000	9.945	-0.055	55980	0.0 %	55980				

LB111758 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Oct 2020 08:23:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB		30 Oct 2020 08:25:42	S	Std
0.2 - 1	50.2	1144	-	0.20007471B	PPB		30 Oct 2020 08:27:59	S	Std
2.5 - 1	52.5	13846	-	2.50007471B	PPB		30 Oct 2020 08:30:16	S	Std
5.0 - 1	55	27291	-	5.00007471B	PPB		30 Oct 2020 08:32:33	S	Std
7.5 - 1	57.5	43139	-	7.50007471B	PPB		30 Oct 2020 08:34:49	S	Std
10.0 - 1	510	55980	-	10.00007471B	PPB		30 Oct 2020 08:42:13	S	Std
ICV80 - 1	ICV80	21470	3.8269	-7471B	PPB		30 Oct 2020 08:44:57	U	SMPL
ICB80 - 1	ICB80	-69	0.0086	-7471B	PPB		30 Oct 2020 08:47:12	U	SMPL
CCV35 - 1	CCV35	27230	4.8480	-7471B	PPB		30 Oct 2020 08:49:27	U	SMPL
CCB35 - 1	CCB35	-74	0.0077	-7471B	PPB		30 Oct 2020 08:51:42	U	SMPL
CRA - 1	CRA	1138	0.2226	-7471B	PPB		30 Oct 2020 08:53:57	U	SMPL
HighStd - 1	HighStd	57006	10.1266	-7471B	PPB		30 Oct 2020 08:56:12	U	SMPL
ChkStd - 1	ChkStd	40992	7.2877	-7471B	PPB		30 Oct 2020 08:58:28	U	SMPL
PB132707BL - 1	PBS	-154	-0.0065	-7471B	PPB		30 Oct 2020 09:00:45	U	SMPL
PB132707BS - 1	LCSS	21342	3.8042	-7471B	PPB		30 Oct 2020 09:03:02	U	SMPL
L4215-01 - 1	AU-05-102820	3018	0.5558	-7471B	PPB		30 Oct 2020 09:05:18	U	SMPL
L4217-07 - 1	HR-01-102820	2699	0.4993	-7471B	PPB		30 Oct 2020 09:07:34	U	SMPL
L4217-07DUP - 1	HR-01-102820DUP	2475	0.4596	-7471B	PPB		30 Oct 2020 09:09:51	U	SMPL
L4217-07MS - 1	HR-01-102820MS	23253	4.1430	-7471B	PPB		30 Oct 2020 09:12:08	U	SMPL
L4217-07MSD - 1	HR-01-102820MSD	23276	4.1471	-7471B	PPB		30 Oct 2020 09:14:23	U	SMPL
CCV36 - 1	CCV36	27186	4.8402	-7471B	PPB		30 Oct 2020 09:16:38	U	SMPL
CCB36 - 1	CCB36	-62	0.0098	-7471B	PPB		30 Oct 2020 09:18:54	U	SMPL
L4227-07 - 1	BU-02-102820	1826	0.3445	-7471B	PPB		30 Oct 2020 09:21:09	U	SMPL
L4531-04 - 1	COMP-A-B-C	327	0.0788	-7471B	PPB		30 Oct 2020 09:23:24	U	SMPL
L4539-02 - 1	JCG-SOIL	1618	0.3077	-7471B	PPB		30 Oct 2020 09:25:39	U	SMPL
L4540-02 - 1	341	3747	0.6851	-7471B	PPB		30 Oct 2020 09:27:55	U	SMPL
L4541-02 - 1	124031	2317	0.4316	-7471B	PPB		30 Oct 2020 09:30:11	U	SMPL
L4541-04 - 1	124024	4780	0.8682	-7471B	PPB		30 Oct 2020 09:32:27	U	SMPL
L4541-06 - 1	124036	735	0.1511	-7471B	PPB		30 Oct 2020 09:34:43	U	SMPL
L4541-08 - 1	20-105	899	0.1802	-7471B	PPB		30 Oct 2020 09:37:00	U	SMPL
L4541-10 - 1	2015	2891	0.5333	-7471B	PPB		30 Oct 2020 09:39:17	U	SMPL
L4547-02 - 1	TT-033-IDWSO-20201027	63	0.0320	-7471B	PPB		30 Oct 2020 09:41:32	U	SMPL
CCV37 - 1	CCV37	27231	4.8482	-7471B	PPB		30 Oct 2020 09:43:48	U	SMPL
CCB37 - 1	CCB37	-32	0.0151	-7471B	PPB		30 Oct 2020 09:46:03	U	SMPL
L4217-07LX5 - 1		521	0.1132	-7471B	PPB		30 Oct 2020 09:48:18	U	SMPL
L4217-07A - 1		25705	4.5777	-7471B	PPB		30 Oct 2020 09:50:34	U	SMPL
CCV38 - 1	CCV38	27307	4.8617	-7471B	PPB		30 Oct 2020 09:52:50	U	SMPL
CCB38 - 1	CCB38	-82	0.0063	-7471B	PPB		30 Oct 2020 09:55:05	U	SMPL