

LB111767


SFAM01.0

INSTRUMENT ID: CV2

Linear

 μ Abs.:

32898



A= 0.0000e+000

B= 3.0046e-004

C= -4.5698e-002

Rho= 0.9994823

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	225	0.000	225					=
0.05	0.050					400					-10
0.20	0.200	0.180	-0.020	752	0.0 %	752					-3
2.50	2.500	2.437	-0.063	8264	0.0 %	8264					0
5.00	5.000	4.990	-0.010	16760	0.0 %	16760					3
7.50	7.500	7.732	0.232	25885	0.0 %	25885					-2
10.0	10.000	9.839	-0.161	32898	0.0 %	32898					

LB111767
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	225	-		0 SFAM01.0	PPB	10/30/2020 14:02	Std
	0.2 S0.2	752	-		0.2 SFAM01.0	PPB	10/30/2020 14:04	Std
	2.5 S2.5	8264	-		2.5 SFAM01.0	PPB	10/30/2020 14:06	Std
	5 S5	16760	-		5 SFAM01.0	PPB	10/30/2020 14:09	Std
	7.5 S7.5	25885	-		7.5 SFAM01.0	PPB	10/30/2020 14:13	Std
	10 S10	32898	-		10 SFAM01.0	PPB	10/30/2020 14:17	Std
ICV	ICV	13801	4.1009 -		SFAM01.0	PPB	10/30/2020 14:20	SMPL
ICB	ICB	32	-0.0361 -		SFAM01.0	PPB	10/30/2020 14:22	SMPL
CCV042	CCV042	16922	5.0387 -		SFAM01.0	PPB	10/30/2020 14:25	SMPL
CCB042	CCB042	40	-0.0337 -		SFAM01.0	PPB	10/30/2020 14:27	SMPL
PBW012	PBW012	157	0.0015 -		SFAM01.0	PPB	10/30/2020 14:29	SMPL
PBW013	PBW013	166	0.0042 -		SFAM01.0	PPB	10/30/2020 14:31	SMPL
PBW014	PBW014	172	0.006 -		SFAM01.0	PPB	10/30/2020 14:34	SMPL
L4485-12	MDL-WATER-05	257	0.0315 -		SFAM01.0	PPB	10/30/2020 14:48	SMPL
L4485-13	MDL-WATER-06	275	0.0369 -		SFAM01.0	PPB	10/30/2020 14:53	SMPL
L4485-14	MDL-WATER-07	297	0.0435 -		SFAM01.0	PPB	10/30/2020 14:55	SMPL
CCV043	CCV043	16905	5.0336 -		SFAM01.0	PPB	10/30/2020 15:00	SMPL
CCB043	CCB043	61	-0.0274 -		SFAM01.0	PPB	10/30/2020 15:03	SMPL