



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111773

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/30/2020	15:01
2	CAL2	1	Water	NA	NA	20.1	6.99	10/30/2020	15:02
3	CAL3	1	Water	NA	NA	20.1	10.03	10/30/2020	15:03
4	ICV	1	Water	NA	NA	20.1	7.01	10/30/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	10/30/2020	15:06
6	L4569-03	1	Water	NA	NA	13.1	7.56	10/30/2020	15:10
7	L4569-03DUP	1	Water	NA	NA	13.2	7.57	10/30/2020	15:11
8	CCV2	1	Water	NA	NA	20.1	12.02	10/30/2020	15:13

LB 111773

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4569 WorkList ID : 143802 Department : Wet-Chemistry Date : 10-30-2020 14:54:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2020	Water	L4569-03 -D	pH	Cool 4 deg C	PSEG01	E12	16122	10/30/2020	9040C

Date/Time 10/30/2020 15:00
Received by: JL Wet
Relinquished by: JL Wet

Date/Time 10/30/2020 16:00
Received by: JL Wet
Relinquished by: JL Wet