



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111782

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/29/2020	09:20
2	L4539-04		1	8	100.00	>212.0	>212.0	10/29/2020	09:45
3	L4539-04DUP		1	8	100.00	>212.0	>212.0	10/29/2020	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111782

WorkList Name : fp102820

WorkList ID : 143692

Department : Wet-Chemistry

Date : 10-28-2020 09:05:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2020	Water	L4539-04	C Flash Point	Cool 4 deg C	PSEG01	D61	FM655	10/28/2020	1010A

Date/Time 10/29/20 08:45
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 10/29/20 10:30
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)