

LB111786


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

35668



A= 0.0000e+000

B= 2.8607e-004

C= -1.8323e-001

Rho= 0.9999828

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	591	0.000	591				
0.05	0.050	0.057	0.007	841	0.0 %	841	0			
0.2	0.200	0.192	-0.008	1313	0.0 %	1313				
2.5	2.500	2.530	0.030	9483	0.0 %	9483				
5.0	5.000	5.005	0.005	18137	0.0 %	18137				
7.5	7.500	7.459	-0.041	26714	0.0 %	26714				
10.0	10.000	10.020	0.020	35668	0.0 %	35668				

LB111786 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Nov 2020 11:24:00

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	591	-	0.0000245.1	PPB	02 Nov 2020 11:38:16	S	Std
0.05 - 1	50.05	841	-	0.0500245.1	PPB	02 Nov 2020 11:40:33	S	Std
0.2 - 1	50.2	1313	-	0.2000245.1	PPB	02 Nov 2020 11:42:50	S	Std
2.5 - 1	52.5	9483	-	2.5000245.1	PPB	02 Nov 2020 11:45:07	S	Std
5.0 - 1	55	18137	-	5.0000245.1	PPB	02 Nov 2020 11:47:23	S	Std
7.5 - 1	57.5	26714	-	7.5000245.1	PPB	02 Nov 2020 11:49:39	S	Std
10.0 - 1	514	35668	-	10.0000245.1	PPB	02 Nov 2020 11:51:56	S	Std
ICV83 - 1	ICV83	15016	4.1125	-245.1	PPB	02 Nov 2020 12:07:04	U	SMPL
ICB83 - 1	ICB83	481	-0.0456	-245.1	PPB	02 Nov 2020 12:09:19	U	SMPL
CCV49 - 1	CCV49	18219	5.0288	-245.1	PPB	02 Nov 2020 12:11:34	U	SMPL
CCB49 - 1	CCB49	544	-0.0276	-245.1	PPB	02 Nov 2020 12:13:49	U	SMPL
CRA - 1	CRA	810	0.0485	-245.1	PPB	02 Nov 2020 12:16:05	U	SMPL
HighStd - 1	HighStd	35422	9.9501	-245.1	PPB	02 Nov 2020 12:18:20	U	SMPL
ChkStd - 1	ChkStd	25123	7.0038	-245.1	PPB	02 Nov 2020 12:20:35	U	SMPL
PB132768BL - 1	PBW	612	-0.0081	-245.1	PPB	02 Nov 2020 12:22:52	U	SMPL
PB132768BS - 1	LCSW	14225	3.8862	-245.1	PPB	02 Nov 2020 12:25:09	U	SMPL
L4533-01 - 1	IB1-TREATED-EFFLU	764	0.0353	-245.1	PPB	02 Nov 2020 12:27:25	U	SMPL
L4533-01DUP - 1	IB1-TREATED-EFFLUDUP	833	0.0551	-245.1	PPB	02 Nov 2020 12:29:42	U	SMPL
L4533-01MS - 1	IB1-TREATED-EFFLUMS	16486	4.5330	-245.1	PPB	02 Nov 2020 12:31:59	U	SMPL
L4533-01MSD - 1	IB1-TREATED-EFFLUMSD	17001	4.6803	-245.1	PPB	02 Nov 2020 12:34:15	U	SMPL
CCV50 - 1	CCV50	18035	4.9761	-245.1	PPB	02 Nov 2020 12:38:46	U	SMPL
CCB50 - 1	CCB50	601	-0.0113	-245.1	PPB	02 Nov 2020 12:41:01	U	SMPL
L4533-02 - 1	IB1-TREATED-EFFLU	418	-0.0636	-245.1	PPB	02 Nov 2020 12:47:37	U	SMPL
L4573-03 - 1	OB-722	1571	0.2662	-245.1	PPB	02 Nov 2020 12:49:52	U	SMPL
L4533-01LX5 - 1		567	-0.0210	-245.1	PPB	02 Nov 2020 12:52:07	U	SMPL
L4533-01A - 1		14434	3.9460	-245.1	PPB	02 Nov 2020 12:54:23	U	SMPL
CCV51 - 1	CCV51	18359	5.0688	-245.1	PPB	02 Nov 2020 12:56:39	U	SMPL
CCB51 - 1	CCB51	519	-0.0348	-245.1	PPB	02 Nov 2020 12:58:55	U	SMPL