

LB 111791


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

38061



A= 0.0000e+000

B= 2.6596e-004

C= -2.5339e-001

Rho= 0.9994811

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	949	0.000	949				
0.2	0.200	0.179	-0.021	1627	0.0 %	1627				
2.5	2.500	2.542	0.042	10510	0.0 %	10510				
5.0	5.000	4.883	-0.117	19312	0.0 %	19312				
7.5	7.500	7.728	0.228	30008	0.0 %	30008				
10.0	10.000	9.869	-0.131	38061	0.0 %	38061				

LB111791 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Nov 2020 13:40:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	949	-	0.0000	7470A	PPB	02 Nov 2020 13:41:39	S	Std
0.2 - 1	50.2	1627	-	0.2000	7470A	PPB	02 Nov 2020 13:43:56	S	Std
2.5 - 1	52.5	10510	-	2.5000	7470A	PPB	02 Nov 2020 13:46:13	S	Std
5.0 - 1	55	19312	-	5.0000	7470A	PPB	02 Nov 2020 13:48:29	S	Std
7.5 - 1	57.5	30008	-	7.5000	7470A	PPB	02 Nov 2020 13:50:46	S	Std
10.0 - 1	510	38061	-	10.0000	7470A	PPB	02 Nov 2020 13:55:44	S	Std
ICV84 - 1	ICV84	15784	3.9445	-	7470A	PPB	02 Nov 2020 14:00:54	U	SMPL
ICB84 - 1	ICB84	786	-0.0443	-	7470A	PPB	02 Nov 2020 14:03:09	U	SMPL
CCV52 - 1	CCV52	19451	4.9198	-	7470A	PPB	02 Nov 2020 14:05:24	U	SMPL
CCB52 - 1	CCB52	769	-0.0489	-	7470A	PPB	02 Nov 2020 14:07:39	U	SMPL
CRA - 1	CRA	1534	0.1546	-	7470A	PPB	02 Nov 2020 14:09:54	U	SMPL
HighStd - 1	HighStd	38880	10.0872	-	7470A	PPB	02 Nov 2020 14:12:09	U	SMPL
ChkStd - 1	ChkStd	26575	6.8145	-	7470A	PPB	02 Nov 2020 14:14:24	U	SMPL
PB132767BL - 1	PBW	1220	0.0711	-	7470A	PPB	02 Nov 2020 14:21:42	U	SMPL
PB132767BS - 1	LCSW	15131	3.7709	-	7470A	PPB	02 Nov 2020 14:23:58	U	SMPL
L4569-03 - 1	16122	757	-0.0521	-	7470A	PPB	02 Nov 2020 14:30:27	U	SMPL
L4569-03DUP - 1	16122DUP	857	-0.0255	-	7470A	PPB	02 Nov 2020 14:35:15	U	SMPL
L4569-03MS - 1	16122MS	13898	3.4429	-	7470A	PPB	02 Nov 2020 14:37:32	U	SMPL
L4569-03MSD - 1	16122MSD	14124	3.5030	-	7470A	PPB	02 Nov 2020 14:39:49	U	SMPL
CCV53 - 1	CCV53	19565	4.9501	-	7470A	PPB	02 Nov 2020 14:42:04	U	SMPL
CCB53 - 1	CCB53	798	-0.0411	-	7470A	PPB	02 Nov 2020 14:44:19	U	SMPL
PB132767TB - 1	PB132759TB	967	0.0038	-	7470A	PPB	02 Nov 2020 14:49:11	U	SMPL
L4569-03LX5 - 1		882	-0.0188	-	7470A	PPB	02 Nov 2020 14:51:26	U	SMPL
L4569-03A - 1		15663	3.9124	-	7470A	PPB	02 Nov 2020 14:53:41	U	SMPL
CCV54 - 1	CCV54	19207	4.8549	-	7470A	PPB	02 Nov 2020 14:55:56	U	SMPL
CCB54 - 1	CCB54	916	-0.0098	-	7470A	PPB	02 Nov 2020 15:00:45	U	SMPL