


LB111800

7471B

INSTRUMENT ID: CV1

μ Abs.:

52840



Linear

A= 0.0000e+000

B= 1.9091e-004

C= -7.1096e-002

Rho= 0.9996831

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.068	-0.068	17	0.000	17				
0.2	0.200	0.168	-0.032	1252	0.0 %	1252				
2.5	2.500	2.609	0.109	14038	0.0 %	14038				
5.0	5.000	5.117	0.117	27174	0.0 %	27174				
7.5	7.500	7.357	-0.143	38910	0.0 %	38910				
10.0	10.000	10.017	0.017	52840	0.0 %	52840				

LB111800 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Nov 2020 07:45:28

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	17	-	0.00007471B	PPB	03 Nov 2020 07:47:58	S	Std
0.2 - 1	50.2	1252	-	0.20007471B	PPB	03 Nov 2020 07:52:47	S	Std
2.5 - 1	52.5	14038	-	2.50007471B	PPB	03 Nov 2020 07:55:04	S	Std
5.0 - 1	55	27174	-	5.00007471B	PPB	03 Nov 2020 07:57:20	S	Std
7.5 - 1	57.5	38910	-	7.50007471B	PPB	03 Nov 2020 07:59:37	S	Std
10.0 - 1	59	52840	-	10.00007471B	PPB	03 Nov 2020 08:01:53	S	Std
ICV85 - 1	ICV85	20656	3.8724	-7471B	PPB	03 Nov 2020 08:04:46	U	SMPL
ICB85 - 1	ICB85	16	-0.0680	-7471B	PPB	03 Nov 2020 08:07:01	U	SMPL
CCV55 - 1	CCV55	27741	5.2250	-7471B	PPB	03 Nov 2020 08:09:16	U	SMPL
CCB55 - 1	CCB55	3	-0.0705	-7471B	PPB	03 Nov 2020 08:11:31	U	SMPL
CRA - 1	CRA	1270	0.1714	-7471B	PPB	03 Nov 2020 08:13:48	U	SMPL
HighStd - 1	HighStd	53073	10.0613	-7471B	PPB	03 Nov 2020 08:16:03	U	SMPL
ChkStd - 1	ChkStd	36885	6.9708	-7471B	PPB	03 Nov 2020 08:18:18	U	SMPL
PB132760BL - 1	PBS	104	-0.0512	-7471B	PPB	03 Nov 2020 08:25:24	U	SMPL
PB132760BS - 1	LCSS	21104	3.9580	-7471B	PPB	03 Nov 2020 08:27:40	U	SMPL
L4222-06 - 1	EO-02-102920	2179	0.3449	-7471B	PPB	03 Nov 2020 08:29:56	U	SMPL
L4555-01 - 1	LCP1-SW-67-0-6	18576	3.4753	-7471B	PPB	03 Nov 2020 08:32:13	U	SMPL
L4555-02 - 1	LCP1-SW-68-0-6	79013	15.0136	-7471B	PPB	03 Nov 2020 08:34:30	U	SMPL
L4555-03 - 1	LCP1-SED-123-0-6	8441	1.5404	-7471B	PPB	03 Nov 2020 08:36:48	U	SMPL
L4555-04 - 1	LCP1-SED-116-0-6	6754	1.2183	-7471B	PPB	03 Nov 2020 08:39:06	U	SMPL
CCV56 - 1	CCV56	27434	5.1664	-7471B	PPB	03 Nov 2020 08:41:22	U	SMPL
CCB56 - 1	CCB56	24	-0.0665	-7471B	PPB	03 Nov 2020 08:43:37	U	SMPL
L4555-05 - 1	LCP1-SED-134-0-6	386360	HIGH	-7471B	PPB	03 Nov 2020 08:45:53	X	
L4555-06 - 1	LCP1-SED-119-0-6	403860	HIGH	-7471B	PPB	03 Nov 2020 08:47:44	X	
L4555-07 - 1	LCP1-SED-129-0-6	317066	60.4612	-7471B	PPB	03 Nov 2020 08:49:38	U	SMPL
L4555-08 - 1	LCP1-SED-132-0-6	368650	70.3093	-7471B	PPB	03 Nov 2020 08:51:57	U	SMPL
L4555-09 - 1	LCP1-SED-118-0-6	313173	HIGH	-7471B	PPB	03 Nov 2020 08:54:25	X	
L4555-10 - 1	LCP1-SED-131-0-6	120759	22.9835	-7471B	PPB	03 Nov 2020 08:56:31	U	SMPL
L4564-01 - 1	FK-GF-03-017-A	9895	1.8180	-7471B	PPB	03 Nov 2020 08:58:51	U	SMPL
L4564-03 - 1	FK-GF-03-017-B	6908	1.2477	-7471B	PPB	03 Nov 2020 09:01:11	U	SMPL
L4565-01 - 1	FK-GF-02-063-A	8108	1.4768	-7471B	PPB	03 Nov 2020 09:03:28	U	SMPL
CCV57 - 1	CCV57	27951	5.2651	-7471B	PPB	03 Nov 2020 09:05:43	U	SMPL
CCB57 - 1	CCB57	-77	-0.0858	-7471B	PPB	03 Nov 2020 09:07:59	U	SMPL
L4565-03 - 1	FK-GF-02-063-B	12073	2.2338	-7471B	PPB	03 Nov 2020 09:15:04	U	SMPL
L4565-03DUP - 1	FK-GF-02-063-BDUP	11498	2.1240	-7471B	PPB	03 Nov 2020 09:17:19	U	SMPL
L4565-03MS - 1	FK-GF-02-063-BMS	32666	6.1653	-7471B	PPB	03 Nov 2020 09:22:11	U	SMPL
L4565-03MSD - 1	FK-GF-02-063-BMSD	31319	5.9081	-7471B	PPB	03 Nov 2020 09:24:26	U	SMPL
L4565-03LX5 - 1		2195	0.3480	-7471B	PPB	03 Nov 2020 09:26:43	U	SMPL
L4565-03A - 1		32389	6.1124	-7471B	PPB	03 Nov 2020 09:29:00	U	SMPL
L4555-02DLX5 - 1		15520	2.8919	-7471B	PPB	03 Nov 2020 09:31:16	U	SMPL
L4555-05DLX10 - 1		59640	11.3150	-7471B	PPB	03 Nov 2020 09:33:34	U	SMPL
L4555-06DLX10 - 1		64269	12.1988	-7471B	PPB	03 Nov 2020 09:35:50	U	SMPL
L4555-07DLX10 - 1		34917	6.5950	-7471B	PPB	03 Nov 2020 09:38:09	U	SMPL
CCV58 - 1	CCV58	27432	5.1661	-7471B	PPB	03 Nov 2020 09:40:27	U	SMPL
CCB58 - 1	CCB58	-18	-0.0745	-7471B	PPB	03 Nov 2020 09:42:44	U	SMPL
L4555-08DLX10 - 1		41987	7.9448	-7471B	PPB	03 Nov 2020 09:45:00	U	SMPL
L4555-09DLX10 - 1		43562	8.2455	-7471B	PPB	03 Nov 2020 09:47:16	U	SMPL
L4555-10DLX5 - 1		24772	4.6582	-7471B	PPB	03 Nov 2020 09:52:07	U	SMPL
L4555-05DL2X25 - 1		24672	4.6391	-7471B	PPB	03 Nov 2020 10:01:07	U	SMPL
L4555-06DL2X25 - 1		26850	5.0549	-7471B	PPB	03 Nov 2020 10:03:22	U	SMPL
CCV59 - 1	CCV59	27744	5.2256	-7471B	PPB	03 Nov 2020 10:05:38	U	SMPL
CCB59 - 1	CCB59	177	-0.0373	-7471B	PPB	03 Nov 2020 10:07:55	U	SMPL