


LB111808

245.1

INSTRUMENT ID: CV1

 μ Abs.:

32355



Linear

A= 0.0000e+000

B= 3.0509e-004

C= -3.3115e-002

Rho= 0.9996462

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	120	0.000	120				
0.05	0.050	0.044	-0.006	254	0.0 %	254	0			
0.2	0.200	0.168	-0.032	660	0.0 %	660				
2.5	2.500	2.434	-0.066	8087	0.0 %	8087				
5.0	5.000	5.096	0.096	16812	0.0 %	16812				
7.5	7.500	7.666	0.166	25234	0.0 %	25234				
10.0	10.000	9.838	-0.162	32355	0.0 %	32355				

LB111803 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Nov 2020 11:46:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	120	-	0.0000	245.1	PPB	03 Nov 2020 11:48:09	S	Std
0.05 - 1	50.05	254	-	0.0500	245.1	PPB	03 Nov 2020 11:50:25	S	Std
0.2 - 1	50.2	660	-	0.2000	245.1	PPB	03 Nov 2020 11:52:42	S	Std
2.5 - 1	52.5	8087	-	2.5000	245.1	PPB	03 Nov 2020 11:54:59	S	Std
5.0 - 1	55	16812	-	5.0000	245.1	PPB	03 Nov 2020 11:57:16	S	Std
7.5 - 1	57.5	25234	-	7.5000	245.1	PPB	03 Nov 2020 11:59:32	S	Std
10.0 - 1	518	32355	-	10.0000	245.1	PPB	03 Nov 2020 12:07:12	S	Std
ICV86 - 1	ICV86	12652	3.8269	-245.1		PPB	03 Nov 2020 12:09:58	U	SMPL
ICB86 - 1	ICB86	156	0.0145	-245.1		PPB	03 Nov 2020 12:12:13	U	SMPL
CCV60 - 1	CCV60	17041	5.1659	-245.1		PPB	03 Nov 2020 12:14:29	U	SMPL
CCB60 - 1	CCB60	157	0.0148	-245.1		PPB	03 Nov 2020 12:16:43	U	SMPL
CRA - 1	CRA	262	0.0468	-245.1		PPB	03 Nov 2020 12:21:29	U	SMPL
HighStd - 1	HighStd	32336	9.8323	-245.1		PPB	03 Nov 2020 12:23:44	U	SMPL
ChkStd - 1	ChkStd	23779	7.2216	-245.1		PPB	03 Nov 2020 12:25:59	U	SMPL
PB132756BL - 1	PBW	81	-0.0084	-245.1		PPB	03 Nov 2020 12:30:50	U	SMPL
PB132756BS - 1	LCSW	12202	3.6896	-245.1		PPB	03 Nov 2020 12:33:05	U	SMPL
L4576-01 - 1	PW-1	-79	-0.0572	-245.1		PPB	03 Nov 2020 12:35:21	U	SMPL
L4576-01DUP - 1	PW-1DUP	-12	-0.0368	-245.1		PPB	03 Nov 2020 12:37:38	U	SMPL
L4576-01MS - 1	PW-1MS	12326	3.7274	-245.1		PPB	03 Nov 2020 12:39:54	U	SMPL
L4576-01MSD - 1	PW-1MSD	12555	3.7973	-245.1		PPB	03 Nov 2020 12:42:11	U	SMPL
L4576-02 - 1	PW-2	-76	-0.0563	-245.1		PPB	03 Nov 2020 12:44:26	U	SMPL
CCV61 - 1	CCV61	16469	4.9914	-245.1		PPB	03 Nov 2020 12:46:42	U	SMPL
CCB61 - 1	CCB61	147	0.0117	-245.1		PPB	03 Nov 2020 12:48:57	U	SMPL
L4576-01LX5 - 1		206	0.0297	-245.1		PPB	03 Nov 2020 12:51:12	U	SMPL
L4576-01A - 1		12497	3.7796	-245.1		PPB	03 Nov 2020 12:53:28	U	SMPL
CCV62 - 1	CCV62	17074	5.1760	-245.1		PPB	03 Nov 2020 12:55:43	U	SMPL
CCB62 - 1	CCB62	171	0.0191	-245.1		PPB	03 Nov 2020 12:57:59	U	SMPL