



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111806

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/03/2020	14:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/03/2020	14:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/03/2020	14:17
4	ICV	1	Water	NA	NA	20.2	7.01	11/03/2020	14:19
5	CCV1	1	Water	NA	NA	20.2	2.02	11/03/2020	14:20
6	L4620-01	1	Water	NA	NA	16.0	5.81	11/03/2020	14:22
7	L4620-01DUP	1	Water	NA	NA	16.1	5.82	11/03/2020	14:23
8	CCV2	1	Water	NA	NA	20.1	12.02	11/03/2020	14:25

11806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4620 WorkList ID : 143865 Department : Wet-Chemistry Date : 11-03-2020 14:03:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2020	Water	L4620-01- <u>B</u>	pH	Cool 4 deg C	STER05	E61	BOTTLE-1	11/02/2020	SM 4500-H B

Date/Time 11-03-2020 14:10
Received by: J. A. (Sul)
Relinquished by: J. A. (Sul)

Date/Time 11-03-2020 15:40
Received by: J. A. (Sul)
Relinquished by: J. A. (Sul)