



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111812

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/03/2020	16:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/03/2020	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/03/2020	16:18
4	ICV	1	Water	NA	NA	20.1	7.02	11/03/2020	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	11/03/2020	16:22
6	L4612-01	1	Water	NA	NA	14.5	6.84	11/03/2020	16:25
7	L4612-01DUP	1	Water	NA	NA	14.6	6.85	11/03/2020	16:26
8	CCV2	1	Water	NA	NA	20.1	12.02	11/03/2020	16:27

103111812

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4612 WorkList ID : 143867 Department : Wet-Chemistry Date : 11-03-2020 16:06:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/03/2020	Water	L4612-01	pH	Cool 4 deg C	PSEG03	E31	265316	11/03/2020	9040C

Date/Time 11-03-2020 16:10
Received by: Jn (wetchem)
Relinquished by: CP (Sml)

Date/Time 11-03-2020 17:14
Received by: CP (Sml)
Relinquished by: Jn (wetchem)