



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111814

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/03/2020	16:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/04/2020	16:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/03/2020	16:18
4	ICV	1	Water	NA	NA	20.1	7.02	11/03/2020	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	11/03/2020	16:22
6	L4611-01	1	Solid	20.03	20	17.2	6.90	11/03/2020	16:45
7	L4611-01DUP	1	Solid	20.04	20	20.04	6.91	11/03/2020	16:46
8	L4621-04	1	Solid	20.02	20	16.6	9.13	11/03/2020	16:50
9	L4621-06	1	Solid	20.03	20	16.6	8.04	11/03/2020	16:51
10	CCV2	1	Water	NA	NA	20.2	12.02	11/03/2020	17:00

WORKLIST(Hardcopy Internal Chain)

111814

WorkList Name : CORROSIONITY L4621 WorkList ID : 143869 Department : Wet-Chemistry Date : 11-03-2020 16:36:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/03/2020	Solid	L4611-01	Corrosivity	Cool 4 deg C	PSEG03		BW-SOIL	11/03/2020	9045D
11/03/2020	Solid	L4621-04	Corrosivity	Cool 4 deg C	ENTE01	E51	PAV-B8-110320-0-10	11/03/2020	9045D
11/03/2020	Solid	L4621-06	Corrosivity	Cool 4 deg C	ENTE01	E51	PAV-B8-110320-10-20	11/03/2020	9045D

Date/Time 11-03-2020 16:40
 Received by: J.D. (Wet-Chem)
 Relinquished by: C.P. (Sml)

Date/Time 11-03-2020 17:45
 Received by: C.P. (Sml)
 Relinquished by: J.D. (Wet-Chem)