

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB111825

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP87313
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB132817BL		1	5.02	50	2.00	0.00	1.90	1.90	0.10	0.00	0.00	11/04/2020	12:05
2	L4600-02		1	5.03	50	2.00	1.90	3.74	1.84	0.16	0.06	4.77	11/04/2020	12:08
3	L4600-02DUP		1	5.05	50	2.00	3.74	5.60	1.86	0.14	0.04	3.17	11/04/2020	12:12
4	L4600-06		1	5.08	50	2.00	5.60	7.42	1.82	0.18	0.08	6.30	11/04/2020	12:15
5	L4600-10		1	5.07	50	2.00	7.42	9.30	1.88	0.12	0.02	1.58	11/04/2020	12:18
6	L4600-14		1	5.09	50	2.00	0.00	1.84	1.84	0.16	0.06	4.72	11/04/2020	12:22
7	L4605-02		1	5.06	50	2.00	1.84	3.70	1.86	0.14	0.04	3.16	11/04/2020	12:25
8	L4611-01		1	5.05	50	2.00	3.70	5.52	1.82	0.18	0.08	6.34	11/04/2020	12:28
9	L4621-04		1	5.04	50	2.00	5.52	7.32	1.80	0.20	0.10	7.94	11/04/2020	12:32
10	L4621-06		1	5.01	50	2.00	7.32	9.18	1.86	0.14	0.04	3.19	11/04/2020	12:35

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume