

LB111827


INSTRUMENT ID: CY1

7471B

Linear

 μ Abs.:

49438



A= 0.0000e+000

B= 2.0162e-004

C= 2.8625e-002

Rho= 0.9996068

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	27	0.000	27				
0.2	0.200	0.221	0.021	956	0.0 %	956				
2.5	2.500	2.535	0.035	12433	0.0 %	12433				
5.0	5.000	4.787	-0.213	23599	0.0 %	23599				
7.5	7.500	7.626	0.126	37681	0.0 %	37681				
10.0	10.000	9.996	-0.004	49438	0.0 %	49438				

Method: 7471B

Operator: Admin

LB111827 INSTRUMENT ID : CV1

Date of Analysis: 04 Nov 2020 13:58:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	7471B	PPB	04 Nov 2020 13:59:37	S	Std
0.2 - 1	50.2	956	-	0.2000	7471B	PPB	04 Nov 2020 14:01:54	S	Std
2.5 - 1	52.1	12433	-	2.5000	7471B	PPB	04 Nov 2020 14:04:11	S	Std
5.0 - 1	55	23599	-	5.0000	7471B	PPB	04 Nov 2020 14:06:28	S	Std
7.5 - 1	57.5	37681	-	7.5000	7471B	PPB	04 Nov 2020 14:08:45	S	Std
10.0 - 1	510	49438	-	10.0000	7471B	PPB	04 Nov 2020 14:13:32	S	Std
ICV88 - 1	ICV88	19480	3.9562	-7471B	PPB	04 Nov 2020 14:16:14	U	SMPL	
ICB88 - 1	ICB88	-64	0.0157	-7471B	PPB	04 Nov 2020 14:18:29	U	SMPL	
CCV70 - 1	CCV70	24413	4.9508	-7471B	PPB	04 Nov 2020 14:20:44	U	SMPL	
CCB70 - 1	CCB70	-106	0.0073	-7471B	PPB	04 Nov 2020 14:23:00	U	SMPL	
CRA - 1	CRA	897	0.2095	-7471B	PPB	04 Nov 2020 14:25:16	U	SMPL	
HighStd - 1	HighStd	49628	10.0348	-7471B	PPB	04 Nov 2020 14:27:30	U	SMPL	
ChkStd - 1	ChkStd	34874	7.0600	-7471B	PPB	04 Nov 2020 14:29:46	U	SMPL	
PB132822BL - 1	PBS	-193	-0.0103	-7471B	PPB	04 Nov 2020 14:32:03	U	SMPL	
PB132822BS - 1	LCSS	18172	3.6925	-7471B	PPB	04 Nov 2020 14:34:20	U	SMPL	
L4591-01 - 1	OK-01-110320	1838	0.3992	-7471B	PPB	04 Nov 2020 14:36:36	U	SMPL	
L4596-01 - 1	TR-05-110320	407	0.1107	-7471B	PPB	04 Nov 2020 14:38:52	U	SMPL	
L4600-01 - 1	CRT-17	3009	0.6353	-7471B	PPB	04 Nov 2020 14:46:24	U	SMPL	
L4600-01DUP - 1	CRT-17DUP	2894	0.6121	-7471B	PPB	04 Nov 2020 14:48:41	U	SMPL	
L4600-01MS - 1	CRT-17MS	20608	4.1837	-7471B	PPB	04 Nov 2020 14:50:56	U	SMPL	
CCV71 - 1	CCV71	22688	4.6030	-7471B	PPB	04 Nov 2020 14:53:11	U	SMPL	
CCB71 - 1	CCB71	-40	0.0206	-7471B	PPB	04 Nov 2020 14:55:27	U	SMPL	
L4600-01MSD - 1	CRT-17MSD	20196	4.1006	-7471B	PPB	04 Nov 2020 14:57:43	U	SMPL	
L4600-05 - 1	CRT-23	2981	0.6297	-7471B	PPB	04 Nov 2020 14:59:58	U	SMPL	
L4600-09 - 1	CRT-19	5844	1.2069	-7471B	PPB	04 Nov 2020 15:02:13	U	SMPL	
L4600-13 - 1	CRT-6	6154	1.2694	-7471B	PPB	04 Nov 2020 15:04:29	U	SMPL	
L4605-01 - 1	CRT-01	2826	0.5984	-7471B	PPB	04 Nov 2020 15:06:45	U	SMPL	
L4606-01 - 1	S-1	675	0.1647	-7471B	PPB	04 Nov 2020 15:09:01	U	SMPL	
L4609-01 - 1	12016	772	0.1843	-7471B	PPB	04 Nov 2020 15:11:17	U	SMPL	
L4613-01 - 1	VNJ-219	2390	0.5105	-7471B	PPB	04 Nov 2020 15:13:34	U	SMPL	
L4613-03 - 1	282	1113	0.2530	-7471B	PPB	04 Nov 2020 15:15:51	U	SMPL	
L4619-01 - 1	TP-2-COMP	630	0.1556	-7471B	PPB	04 Nov 2020 15:18:07	U	SMPL	
CCV72 - 1	CCV72	24138	4.8954	-7471B	PPB	04 Nov 2020 15:20:22	U	SMPL	
CCB72 - 1	CCB72	-33	0.0220	-7471B	PPB	04 Nov 2020 15:22:38	U	SMPL	
L4619-04 - 1	TP-1-3-4-COMP	564	0.1423	-7471B	PPB	04 Nov 2020 15:24:54	U	SMPL	
L4621-03 - 1	PAV-B8-110320-0-10	22487	4.5625	-7471B	PPB	04 Nov 2020 15:27:10	U	SMPL	
L4621-05 - 1	PAV-B8-110320-10-20	1867	0.4051	-7471B	PPB	04 Nov 2020 15:29:25	U	SMPL	
L4600-01LX5 - 1		567	0.1429	-7471B	PPB	04 Nov 2020 15:31:41	U	SMPL	
L4600-01A - 1		22049	4.4742	-7471B	PPB	04 Nov 2020 15:33:57	U	SMPL	
CCV73 - 1	CCV73	23963	4.8601	-7471B	PPB	04 Nov 2020 15:36:13	U	SMPL	
CCB73 - 1	CCB73	-48	0.0189	-7471B	PPB	04 Nov 2020 15:38:30	U	SMPL	