



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111828

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/04/2020	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/04/2020	14:11
3	CAL3	1	Water	NA	NA	20.1	10.04	11/04/2020	14:12
4	ICV	1	Water	NA	NA	20.1	7.01	11/04/2020	14:15
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2020	14:19
6	L4624-02	1	Solid	20.03	20	16.7	5.43	11/04/2020	14:22
7	L4624-02DUP	1	Solid	20.04	20	16.8	5.44	11/04/2020	14:23
8	L4624-05	1	Solid	20.02	20	16.5	5.10	11/04/2020	14:25
9	L4624-08	1	Solid	20.01	20	16.4	5.28	11/04/2020	14:26
10	CCV2	1	Water	NA	NA	20.1	12.02	11/04/2020	14:28

111828

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4624 WorkList ID : 143899 Department : Wet-Chemistry Date : 11-04-2020 13:54:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2020	Solid	L4624-02	pH	Cool 4 deg C	PSEG03	E32	COMP-1	11/04/2020	9045D
11/04/2020	Solid	L4624-05	pH	Cool 4 deg C	PSEG03	E32	COMP-2	11/04/2020	9045D
11/04/2020	Solid	L4624-08	pH	Cool 4 deg C	PSEG03	E32	COMP-3	11/04/2020	9045D

Date/Time 11-04-2020 14:00
Received by: JP Wet Chem
Relinquished by: KS (Wet Chem)

Date/Time 11-04-2020 15:10
Received by: KS (Wet Chem)
Relinquished by: JP Wet Chem