



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111829

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/04/2020	14:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/04/2020	14:11
3	CAL3	1	Water	NA	NA	20.1	10.04	11/04/2020	14:12
4	ICV	1	Water	NA	NA	20.1	7.01	11/04/2020	14:15
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2020	14:19
6	L4624-10	1	Water	NA	NA	15.5	5.49	11/04/2020	14:30
7	L4624-10DUP	1	Water	NA	NA	15.6	5.50	11/04/2020	14:31
8	CCV2	1	Water	NA	NA	20.2	12.02	11/04/2020	14:32

MS 111829

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4624 W WorkList ID : 143900 Department : Wet-Chemistry Date : 11-04-2020 13:55:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2020	Water	L4624-10	pH	Cool 4 deg C	PSEG03	E32	32733	11/04/2020	9040C

Date/Time 11-04-2020 14:00
Received by: 70/Wet-Chem
Relinquished by: KS (Wetchem)

Date/Time 11-04-2020 15:10
Received by: KS (Wetchem)
Relinquished by: 70/Wet-Chem