



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111834

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/04/2020	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	11/04/2020	15:11
3	CAL3	1	Water	NA	NA	20.2	10.02	11/04/2020	15:12
4	ICV	1	Water	NA	NA	20.2	7.01	11/04/2020	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2020	15:16
6	L4631-04	1	Solid	20.03	20	17.5	8.21	11/04/2020	15:19
7	L4631-04DUP	1	Solid	20.04	20	17.6	8.22	11/04/2020	15:20
8	L4631-06	1	Solid	20.03	20	17.6	8.56	11/04/2020	15:22
9	CCV2	1	Water	NA	NA	20.2	12.02	11/04/2020	15:25

WORKLIST(Hardcopy Internal Chain)

VB 111834

WorkList Name : CORROSIVITY L4625 WorkList ID : 143903 Department : Wet-Chemistry Date : 11-04-2020 13:57:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2020	Solid	L4631-04	Corrosivity	Cool 4 deg C	ENTE01	E61	PAV-B11-110420-0-10	11/04/2020	9045D
11/04/2020	Solid	L4631-06	Corrosivity	Cool 4 deg C	ENTE01	E61	PAV-B11-110420-10-15.25	11/04/2020	9045D

Date/Time 11-04-2020 15:00
 Received by: JP (Wet Chem)
 Relinquished by: CP (Sm1)

Date/Time 11-04-2020 14:25
 Received by: CP (Sm1)
 Relinquished by: JP (Wet Chem)