



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111855

Slope : 97.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
TEST PAPERS,PH 0-14,.5 SENSI,100PK	W2692
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	16.8	4.01	11/05/2020	10:18
2	CAL2	1	Water	NA	NA	16.8	7.01	11/05/2020	10:19
3	CAL3	1	Water	NA	NA	16.8	10.02	11/05/2020	10:20
4	ICV	1	Water	NA	NA	16.8	6.99	11/05/2020	10:22
5	CCV1	1	Water	NA	NA	16.7	2.01	11/05/2020	10:24
6	L4642-01	1	Water	NA	NA	16.3	6.49	11/05/2020	10:26
7	L4650-01	1	Water	NA	NA	16.9	4.73	11/05/2020	10:28
8	L4650-02	1	Water	NA	NA	16.8	4.82	11/05/2020	10:30
9	L4650-03	1	Water	NA	NA	16.8	4.80	11/05/2020	10:32
10	L4652-01	1	Water	NA	NA	17.3	4.90	11/05/2020	10:34
11	L4652-01DUP	1	Water	NA	NA	17.3	4.89	11/05/2020	10:36
12	CCV2	1	Water	NA	NA	16.9	12.2	11/05/2020	10:38