


LB111860

7470A

INSTRUMENT ID: CY1

 μ Abs.:

38433



Linear

A= 0.0000e+000

B= 2.6077e-004

C= 6.5195e-002

Rho= 0.9995053

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.095	0.095	114	0.000	114				
0.2	0.200	0.201	0.001	520	0.0 %	520				
2.5	2.500	2.353	-0.147	8773	0.0 %	8773				
5.0	5.000	5.127	0.127	19410	0.0 %	19410				
7.5	7.500	7.337	-0.163	27887	0.0 %	27887				
10.0	10.000	10.087	0.087	38433	0.0 %	38433				

LB111860 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2020 10:04:11

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	114	-	0.00007470A	PPB	06 Nov 2020 10:37:05	S	Std
0.2 - 1	50.2	520	-	0.20007470A	PPB	06 Nov 2020 10:39:23	S	Std
2.5 - 1	50.1	8773	-	2.50007470A	PPB	06 Nov 2020 10:41:39	S	Std
5.0 - 1	55	19410	-	5.00007470A	PPB	06 Nov 2020 10:43:56	S	Std
7.5 - 1	57.1	27887	-	7.50007470A	PPB	06 Nov 2020 10:46:13	S	Std
10.0 - 1	50	38433	-	10.00007470A	PPB	06 Nov 2020 10:51:10	S	Std
ICV91 - 1	ICV91	14497	3.8456	-7470A	PPB	06 Nov 2020 10:53:50	U	SMPL
ICB91 - 1	ICB91	42	0.0761	-7470A	PPB	06 Nov 2020 10:56:05	U	SMPL
CCV83 - 1	CCV83	19433	5.1327	-7470A	PPB	06 Nov 2020 10:58:20	U	SMPL
CCB83 - 1	CCB83	35	0.0743	-7470A	PPB	06 Nov 2020 11:00:35	U	SMPL
CRA - 1	CRA	486	0.1919	-7470A	PPB	06 Nov 2020 11:02:50	U	SMPL
HighStd - 1	HighStd	38511	10.1077	-7470A	PPB	06 Nov 2020 11:05:05	U	SMPL
ChkStd - 1	ChkStd	26640	7.0121	-7470A	PPB	06 Nov 2020 11:07:20	U	SMPL
PB132851BL - 1	PBW	-131	0.0310	-7470A	PPB	06 Nov 2020 11:09:37	U	SMPL
PB132851BS - 1	LCSW	13821	3.6693	-7470A	PPB	06 Nov 2020 11:11:53	U	SMPL
L4624-02 - 1	COMP-1	68	0.0829	-7470A	PPB	06 Nov 2020 11:14:09	U	SMPL
L4624-05 - 1	COMP-2	188	0.1142	-7470A	PPB	06 Nov 2020 11:16:25	U	SMPL
L4624-08 - 1	COMP-3	171	0.1098	-7470A	PPB	06 Nov 2020 11:18:42	U	SMPL
L4625-02 - 1	CONTAINER-001-A-B-COMP	27	0.0722	-7470A	PPB	06 Nov 2020 11:20:59	U	SMPL
L4625-02DUP - 1	CONTAINER-001-A-B-COMPDUP	106	0.0928	-7470A	PPB	06 Nov 2020 11:23:14	U	SMPL
CCV84 - 1	CCV84	19631	5.1844	-7470A	PPB	06 Nov 2020 11:25:29	U	SMPL
CCB84 - 1	CCB84	52	0.0788	-7470A	PPB	06 Nov 2020 11:27:45	U	SMPL
L4625-02MS - 1	CONTAINER-001-A-B-COMPMS	14274	3.7874	-7470A	PPB	06 Nov 2020 11:30:00	U	SMPL
L4625-02MSD - 1	CONTAINER-001-A-B-COMPMSD	14042	3.7269	-7470A	PPB	06 Nov 2020 11:32:15	U	SMPL
L4625-13 - 1	CONTAINER-002-A-B-COMP	-65	0.0482	-7470A	PPB	06 Nov 2020 11:34:30	U	SMPL
PB1032851TB - 1	PB132823TB	71	0.0837	-7470A	PPB	06 Nov 2020 11:36:46	U	SMPL
L4625-02LX5 - 1		73	0.0842	-7470A	PPB	06 Nov 2020 11:39:02	U	SMPL
L4625-02A - 1		14989	3.9739	-7470A	PPB	06 Nov 2020 11:41:18	U	SMPL
CCV85 - 1	CCV85	19424	5.1304	-7470A	PPB	06 Nov 2020 11:43:34	U	SMPL
CCB85 - 1	CCB85	21	0.0707	-7470A	PPB	06 Nov 2020 11:45:52	U	SMPL