



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111863

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/06/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.01	11/06/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/06/2020	10:17
4	ICV	1	Water	NA	NA	20.1	6.98	11/06/2020	10:19
5	CCV1	1	Water	NA	NA	20.2	2.01	11/06/2020	10:20
6	L4651-02	1	Water	NA	NA	18.2	10.14	11/06/2020	10:21
7	L4651-02DUP	1	Water	NA	NA	18.3	10.15	11/06/2020	10:22
8	CCV2	1	Water	NA	NA	20.2	12.02	11/06/2020	10:23

111863

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L5651 w WorkList ID : 143966 Department : Wet-Chemistry Date : 11-06-2020 09:51:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/04/2020	Water	L4651-02	Corrosivity	Cool 4 deg C	ARCA05	E31	PS-2020-WWW-000-03	11/04/2020	9040C

Date/Time 11-06-2020 10:00
Received by: [Signature]
Relinquished by: KS (wetchem)

Date/Time 11-06-2020 12:20
Received by: KS (wetchem)
Relinquished by: [Signature]