



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111864

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/06/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.01	11/06/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/06/2020	10:17
4	ICV	1	Water	NA	NA	20.1	6.98	11/06/2020	10:19
5	CCV1	1	Water	NA	NA	20.2	2.01	11/06/2020	10:20
6	L4636-06	1	Water	NA	NA	18.3	4.94	11/06/2020	10:25
7	L4636-06DUP	1	Water	NA	NA	18.4	4.95	11/06/2020	10:26
8	L4636-07	1	Water	NA	NA	18.3	6.71	11/06/2020	10:29
9	L4636-08	1	Water	NA	NA	18.4	5.08	11/06/2020	10:30
10	L4636-11	1	Water	NA	NA	18.6	6.57	11/06/2020	10:32
11	L4636-12	1	Water	NA	NA	18.4	6.43	11/06/2020	10:33
12	L4636-13	1	Water	NA	NA	18.3	6.64	11/06/2020	10:35
13	L4636-14	1	Water	NA	NA	18.2	7.20	11/06/2020	10:36
14	L4636-15	1	Water	NA	NA	18.4	7.67	11/06/2020	10:39
15	L4636-16	1	Water	NA	NA	18.5	7.36	11/06/2020	10:40
16	CCV2	1	Water	NA	NA	20.2	12.02	11/06/2020	10:42
17	L4636-17	1	Water	NA	NA	18.6	7.13	11/06/2020	10:45
18	L4637-03	1	Water	NA	NA	18.4	7.50	11/06/2020	10:46
19	L4638-01	1	Water	NA	NA	18.4	5.86	11/06/2020	10:48
20	L4639-01	1	Water	NA	NA	18.5	6.71	11/06/2020	10:49

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.2	2.01	11/06/2020	10:50

WORKLIST(Hardcopy Internal Chain)

WB 111864

WorkList Name : pH 14656

WorkList ID : 143967

Department : Wet-Chemistry

Date : 11-06-2020 09:52:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2020	Water	L4636-06	pH	Cool 4 deg C	PSEG03	E31	31297	11/05/2020	9040C
11/05/2020	Water	L4636-07	pH	Cool 4 deg C	PSEG03	E31	31299	11/05/2020	9040C
11/05/2020	Water	L4636-08	pH	Cool 4 deg C	PSEG03	E31	31307	11/05/2020	9040C
11/05/2020	Water	L4636-11	pH	Cool 4 deg C	PSEG03	E31	31311	11/05/2020	9040C
11/05/2020	Water	L4636-12	pH	Cool 4 deg C	PSEG03	E31	31312	11/05/2020	9040C
11/05/2020	Water	L4636-13	pH	Cool 4 deg C	PSEG03	E31	31313	11/05/2020	9040C
11/05/2020	Water	L4636-14	pH	Cool 4 deg C	PSEG03	E31	31314	11/05/2020	9040C
11/05/2020	Water	L4636-15	pH	Cool 4 deg C	PSEG03	E31	31315	11/05/2020	9040C
11/05/2020	Water	L4636-16	pH	Cool 4 deg C	PSEG03	E31	31316	11/05/2020	9040C
11/05/2020	Water	L4636-17	pH	Cool 4 deg C	PSEG03	E31	31317	11/05/2020	9040C
11/05/2020	Water	L4637-03	pH	Cool 4 deg C	PSEG03	E51	TRE-1117	11/05/2020	9040C
11/05/2020	Water	L4638-01	pH	Cool 4 deg C	PSEG03	E41	31302	11/05/2020	9040C
11/05/2020	Water	L4639-01	pH	Cool 4 deg C	PSEG03	E52	NWB-1551	11/05/2020	9040C

Date/Time 11-06-2020 10:00
 Received by: Jg / wet Chem
 Relinquished by: KS (Wet Chem)

Date/Time 11-06-2020 12:35
 Received by: KS (Wet Chem)
 Relinquished by: Jg (Wet Chem)