



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111865

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	11/06/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.01	11/06/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/06/2020	10:17
4	ICV	1	Water	NA	NA	20.1	6.98	11/06/2020	10:19
5	CCV1	1	Water	NA	NA	20.2	2.01	11/06/2020	10:20
6	L4636-01	1	Solid	20.03	20	18.2	8.36	11/06/2020	11:00
7	L4636-02	1	Solid	20.04	20	18.2	7.86	11/06/2020	11:03
8	L4636-03	1	Solid	20.02	20	18.4	7.57	11/06/2020	11:05
9	L4637-01	1	Solid	20.03	20	20.5	8.68	11/06/2020	11:06
10	L4637-01DUP	1	Solid	20.04	20	20.6	8.69	11/06/2020	11:07
11	CCV2	1	Water	NA	NA	20.1	12.02	11/06/2020	11:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L4658

WorkList ID : 143968

Department : Wet-Chemistry

Date : 11-06-2020 09:53:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2020	Solid	L4636-01	pH	Cool 4 deg C	PSEG03	E31	31275	11/05/2020	9045D
11/05/2020	Solid	L4636-02	pH	Cool 4 deg C	PSEG03	E31	31306	11/05/2020	9045D
11/05/2020	Solid	L4636-03	pH	Cool 4 deg C	PSEG03	E31	31308	11/05/2020	9045D
11/05/2020	Solid	L4637-01	pH	Cool 4 deg C	PSEG03	E51	TRE-1059	11/05/2020	9045D

Date/Time

Received by:

Relinquished by:

11.06.2020 10:00

20 (Wetchem)

KS (Wetchem)

Date/Time

Received by:

Relinquished by:

11.06.2020 12:25

KS (Wetchem)

20 (Wetchem)