



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111874

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/06/2020	11:20
2	L4624-10		1	8	100.00	>212.0	>212.0	11/06/2020	11:45
3	L4624-10DUP		1	8	100.00	>212.0	>212.0	11/06/2020	12:10
4	L4636-09		1	9	100.00	>212.0	>212.0	11/06/2020	12:40
5	L4636-18		1	8	100.00	>212.0	>212.0	11/06/2020	13:05
6	L4637-03		1	9	100.00	>212.0	>212.0	11/06/2020	13:35
7	L4638-01		1	9	100.00	>212.0	>212.0	11/06/2020	13:58
8	L4639-01		1	8	100.00	>212.0	>212.0	11/06/2020	14:25
9	L4651-02		1	9	100.00	>212.0	>212.0	11/06/2020	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111874

WorkList Name : fp 11620

WorkList ID : 143963

Department : Wet-Chemistry

Date : 11-06-2020 09:11:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/11/2020	Water	L4624-10	Flash Point	Cool 4 deg C	PSEG03	E32	32733	11/04/2020	1010A
11/19/2020	Water	L4636-09	Flash Point	Cool 4 deg C	PSEG03	E31	DRUM-COMP-1-AU	11/05/2020	1010A
11/19/2020	Water	L4636-18	Flash Point	Cool 4 deg C	PSEG03	E31	TOTE-COMP-AU	11/05/2020	1010A
11/19/2020	Water	L4637-03	Flash Point	Cool 4 deg C	PSEG03	E51	TRE-1117	11/05/2020	1010A
11/19/2020	Water	L4638-01	Flash Point	Cool 4 deg C	PSEG03	E41	31302	11/05/2020	1010A
11/19/2020	Water	L4639-01	Flash Point	Cool 4 deg C	PSEG03	E52	NWB-1551	11/05/2020	1010A
11/12/2020	Water	L4651-02	Flash Point	Cool 4 deg C	ARCA05	E31	PS-2020-WW-000-03	11/04/2020	1010A

Date/Time 11/06/2020 11:00
 Received by: KS (Wetchem)
 Relinquished by: RJ (Ext-196)

Date/Time 11/06/2020 14:55
 Received by: RJ (Ext-196)
 Relinquished by: KS (Wetchem)