


LB111878
INSTRUMENT ID: CX1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	159	0.000	159				
0.2	0.200	0.218	0.018	1283	0.0 %	1283				
2.5	2.500	2.408	-0.092	12123	0.0 %	12123				
5.0	5.000	5.080	0.080	25339	0.0 %	25339				
7.5	7.500	7.598	0.098	37801	0.0 %	37801				
10.0	10.000	9.906	-0.094	49217	0.0 %	49217				

LB111878 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Nov 2020 15:20:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.00007471B	PPB		06 Nov 2020 15:21:23	S	Std
0.2 - 1	50.2	1283	-	0.20007471B	PPB		06 Nov 2020 15:23:40	S	Std
2.5 - 1	52.5	12123	-	2.50007471B	PPB		06 Nov 2020 15:25:58	S	Std
5.0 - 1	55	25339	-	5.00007471B	PPB		06 Nov 2020 15:28:15	S	Std
7.5 - 1	57.5	37801	-	7.50007471B	PPB		06 Nov 2020 15:30:31	S	Std
10.0 - 1	59.5	49217	-	10.00007471B	PPB		06 Nov 2020 15:35:26	S	Std
ICV92 - 1	ICV92	18503	3.6980	-7471B	PPB		06 Nov 2020 15:38:05	U	SMPL
ICB92 - 1	ICB92	5	-0.0407	-7471B	PPB		06 Nov 2020 15:40:20	U	SMPL
CCV88 - 1	CCV88	25079	5.0270	-7471B	PPB		06 Nov 2020 15:42:36	U	SMPL
CCB88 - 1	CCB88	11	-0.0395	-7471B	PPB		06 Nov 2020 15:44:51	U	SMPL
CRA - 1	CRA	1242	0.2093	-7471B	PPB		06 Nov 2020 15:47:07	U	SMPL
HighStd	HighStd	49241	9.9105	-7471B	PPB		06 Nov 2020 15:49:22	U	SMPL
ChkStd - 1	ChkStd	34689	6.9693	-7471B	PPB		06 Nov 2020 15:51:37	U	SMPL
PB132866BL - 1	PBS	-11	-0.0440	-7471B	PPB		06 Nov 2020 15:53:54	U	SMPL
PB132866BS - 1	LCSS	20058	4.0122	-7471B	PPB		06 Nov 2020 15:56:11	U	SMPL
L4635-01 - 1	356	1184	0.1976	-7471B	PPB		06 Nov 2020 15:58:27	U	SMPL
L4636-04 - 1	DRUM-COMP-2-AU	708	0.1014	-7471B	PPB		06 Nov 2020 16:00:44	U	SMPL
L4641-01 - 1	P0102-CAS001-0002-01	5866	1.1439	-7471B	PPB		06 Nov 2020 16:03:00	U	SMPL
L4641-02 - 1	P0102-CAS001-0204-01	3628	0.6915	-7471B	PPB		06 Nov 2020 16:05:17	U	SMPL
L4641-03 - 1	P0102-CAS001-0406-01	3809	0.7281	-7471B	PPB		06 Nov 2020 16:07:33	U	SMPL
CCV89 - 1	CCV89	24798	4.9702	-7471B	PPB		06 Nov 2020 16:09:48	U	SMPL
CCB89 - 1	CCB89	13	-0.0391	-7471B	PPB		06 Nov 2020 16:12:03	U	SMPL
L4641-04 - 1	P0102-CAS002-0002-01	168691	34.0528	-7471B	PPB		06 Nov 2020 16:14:18	U	SMPL
L4641-05 - 1	P0102-CAS002-0204-01	83292	16.7926	-7471B	PPB		06 Nov 2020 16:16:34	U	SMPL
L4641-06 - 1	P0102-CAS002-0406-01	28076	5.6328	-7471B	PPB		06 Nov 2020 16:18:52	U	SMPL
L4641-07 - 1	P0122-CAS001-0002-01	89359	18.0188	-7471B	PPB		06 Nov 2020 16:21:11	U	SMPL
L4641-07DUP - 1	P0122-CAS001-0002-01DUP	92424	18.6383	-7471B	PPB		06 Nov 2020 16:23:28	U	SMPL
L4641-08 - 1	L4641-07MS	114575	23.1153	-7471B	PPB		06 Nov 2020 16:30:41	U	SMPL
L4641-09 - 1	L4641-07MSD	111384	22.4703	-7471B	PPB		06 Nov 2020 16:32:59	U	SMPL
L4641-10 - 1	P0122-CAS001-0002-02	79268	15.9793	-7471B	PPB		06 Nov 2020 16:35:19	U	SMPL
L4641-11 - 1	P0122-CAS001-0204-01	66509	13.4006	-7471B	PPB		06 Nov 2020 16:37:39	U	SMPL
L4641-12 - 1	P0122-CAS001-0406-01	73501	14.8137	-7471B	PPB		06 Nov 2020 16:39:57	U	SMPL
CCV90 - 1	CCV90	24616	4.9335	-7471B	PPB		06 Nov 2020 16:42:15	U	SMPL
CCB90 - 1	CCB90	-40	-0.0498	-7471B	PPB		06 Nov 2020 16:44:33	U	SMPL
L4653-01 - 1	TP-7-COMP	1251	0.2111	-7471B	PPB		06 Nov 2020 16:46:50	U	SMPL
L4656-01 - 1	SR-L14-L13-NORTH	13743	2.7359	-7471B	PPB		06 Nov 2020 16:49:05	U	SMPL
L4656-08 - 1	SR-L14-L13-SOUTH	9677	1.9141	-7471B	PPB		06 Nov 2020 16:51:21	U	SMPL
L4657-02 - 1	PAV-B49-110520-0-10	14585	2.9061	-7471B	PPB		06 Nov 2020 16:53:36	U	SMPL
L4657-03 - 1	PAV-B49-110520-10-18	5800	1.1305	-7471B	PPB		06 Nov 2020 16:55:52	U	SMPL
L4641-07LX5 - 1		15094	3.0090	-7471B	PPB		06 Nov 2020 16:58:08	U	SMPL
L4641-07A - 1		114901	23.1812	-7471B	PPB		06 Nov 2020 17:00:24	U	SMPL
L4641-04DLX10 - 1		15145	3.0193	-7471B	PPB		06 Nov 2020 17:05:45	U	SMPL
L4641-05DLX5 - 1		14032	2.7943	-7471B	PPB		06 Nov 2020 17:08:01	U	SMPL
CCV91 - 1	CCV91	24802	4.9711	-7471B	PPB		06 Nov 2020 17:10:18	U	SMPL
CCB91 - 1	CCB91	4	-0.0409	-7471B	PPB		06 Nov 2020 17:12:34	U	SMPL
L4641-07DLX5 - 1		14919	2.9736	-7471B	PPB		06 Nov 2020 17:14:50	U	SMPL
L4641-07DUPDLX5 - 1		15751	3.1417	-7471B	PPB		06 Nov 2020 17:17:06	U	SMPL
L4641-08DLX5 - 1		19633	3.9263	-7471B	PPB		06 Nov 2020 17:19:22	U	SMPL
L4641-09DLX5 - 1		20398	4.0810	-7471B	PPB		06 Nov 2020 17:21:39	U	SMPL
L4641-07LDLX25 - 1		3680	0.7020	-7471B	PPB		06 Nov 2020 17:23:55	U	SMPL
L4641-07ADLX5 - 1		22823	4.5711	-7471B	PPB		06 Nov 2020 17:26:12	U	SMPL
L4641-10DLX5 - 1		12917	2.5690	-7471B	PPB		06 Nov 2020 17:28:28	U	SMPL
L4641-11DLX5 - 1		10838	2.1488	-7471B	PPB		06 Nov 2020 17:30:45	U	SMPL
L4641-12DLX5 - 1		14521	2.8931	-7471B	PPB		06 Nov 2020 17:33:03	U	SMPL
CCV92 - 1	CCV92	24791	4.9688	-7471B	PPB		06 Nov 2020 17:35:20	U	SMPL
CCB92 - 1	CCB92	28	-0.0361	-7471B	PPB		06 Nov 2020 17:37:36	U	SMPL