



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111880

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/06/2020	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	11/06/2020	16:26
3	CAL3	1	Water	NA	NA	20.3	10.03	11/06/2020	16:27
4	ICV	1	Water	NA	NA	20.3	7.01	11/06/2020	16:29
5	CCV1	1	Water	NA	NA	20.3	2.01	11/06/2020	16:30
6	L4660-01	1	Solid	20.03	20	20.3	8.85	11/06/2020	16:32
7	L4660-01DUP	1	Solid	20.04	20	20.4	8.86	11/06/2020	16:33
8	CCV2	1	Water	NA	NA	20.2	12.02	11/06/2020	16:35

WORKLIST(Hardcopy Internal Chain)

43111880

WorkList Name : PH L4660 WorkList ID : 143982 Department : Wet-Chemistry Date : 11-06-2020 16:10:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2020	Solid	L4660-01	pH	Cool 4 deg C	PSEG03	E61	CHRT-26601	11/06/2020	9045D

Date/Time 11-06-2020 16:10
 Received by: To wet Chem
 Relinquished by: AP

Date/Time 11-06-2020 17:20
 Received by: AP
 Relinquished by: To wet Chem