



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111886

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/09/2020	11:35
2	CAL2	1	Water	NA	NA	20.1	7.00	11/09/2020	11:36
3	CAL3	1	Water	NA	NA	20.1	10.03	11/09/2020	11:37
4	ICV	1	Water	NA	NA	20.2	7.02	11/09/2020	11:40
5	CCV1	1	Water	NA	NA	20.2	2.01	11/09/2020	11:42
6	L4675-01	1	Water	NA	NA	19.6	5.93	11/09/2020	11:45
7	L4675-01DUP	1	Water	NA	NA	19.7	5.94	11/09/2020	11:46
8	CCV2	1	Water	NA	NA	20.2	12.02	11/09/2020	11:48

2B 111886

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH L4675	WorkList ID :	144005	Department :	Wet-Chemistry	Date :	11-09-2020 11:26:04		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2020	Water	L4675-01	pH	Cool 4 deg C	STER01	E41	POOL-WATER	11/06/2020	9040C

Date/Time 11-09-2020 11:30
Received by: JP Wet Chem
Relinquished by: JP Wet Chem

Date/Time 11-09-2020 11:30
Received by: JP Wet Chem
Relinquished by: JP Wet Chem