


LB111887

ISM02.4

INSTRUMENT ID: CY1



A= 0.0000e+000

B= 2.5032e-004

C= -1.2727e-002

Rho= 0.9999536

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.008	-0.008	17	0.000	17					-4
0.05	0.050	0.048	-0.002	244	0.0 %	244					-17
0.20	0.200	0.167	-0.033	717	0.0 %	717					0
2.50	2.500	2.508	0.008	10069	0.0 %	10069					0
5.00	5.000	5.080	0.080	20344	0.0 %	20344					0
7.50	7.500	7.492	-0.008	29980	0.0 %	29980					0
10.0	10.000	9.964	-0.036	39857	0.0 %	39857					0

LB111887

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	17	-	0	ISM02.4	PPB	11/9/2020 13:14	Std
	0.05 S0.05	244	-	0.05	ISM02.4	PPB	11/9/2020 13:16	Std
	0.2 S0.2	717	-	0.2	ISM02.4	PPB	11/9/2020 13:19	Std
	2.5 S2.5	10069	-	2.5	ISM02.4	PPB	11/9/2020 13:21	Std
	5 S5.0	20344	-	5	ISM02.4	PPB	11/9/2020 13:23	Std
	7.5 S7.5	29980	-	7.5	ISM02.4	PPB	11/9/2020 13:25	Std
	10 S10.0	39857	-	10	ISM02.4	PPB	11/9/2020 13:30	Std
ICV	ICV	15826	3.9488 -		ISM02.4	PPB	11/9/2020 13:35	SMPL
ICB	ICB	64	0.0033 -		ISM02.4	PPB	11/9/2020 13:43	SMPL
CCV097	CCV097	20132	5.0267 -		ISM02.4	PPB	11/9/2020 13:45	SMPL
CCB097	CCB097	-14	-0.0162 -		ISM02.4	PPB	11/9/2020 13:47	SMPL
PB132902BL	PBW902	46	-0.0012 -		ISM02.4	PPB	11/9/2020 13:52	SMPL
L4626-01	MEW0T4	975	0.2313 -		ISM02.4	PPB	11/9/2020 13:54	SMPL
L4626-02	MEW0T5	772	0.1805 -		ISM02.4	PPB	11/9/2020 13:56	SMPL
L4626-03	MEW0T7	669	0.1547 -		ISM02.4	PPB	11/9/2020 13:59	SMPL
L4626-04	MEW0T7D	620	0.1425 -		ISM02.4	PPB	11/9/2020 14:01	SMPL
L4626-05	MEW0T7S	4334	1.0721 -		ISM02.4	PPB	11/9/2020 14:03	SMPL
CCV098	CCV098	19838	4.9531 -		ISM02.4	PPB	11/9/2020 14:05	SMPL
CCB098	CCB098	-43	-0.0235 -		ISM02.4	PPB	11/9/2020 14:08	SMPL