



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/10/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 11/09/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 11/10/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111888

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4559-11	200FB-022-0002	1	1.18	9.58	8.76	90.2	
L4559-12	200FB-022-0204	2	1.16	9.83	9.00	90.4	
L4594-01	OR-03-110920	18	1.15	9.66	9.14	93.9	
L4594-02	OR-03-110920-EPH-2	19	1.11	9.85	9.18	92.3	
L4656-20	SR-L14-L13-NORTH	3	1.12	9.6	8.09	82.2	
L4656-21	SR-L14-L13-SOUTH	4	1.16	9.85	8.37	83.0	
L4657-13	PAV-B49-110520-0-1	5	1.1	9.52	8.87	92.3	
L4657-14	PAV-B49-110520-17-18	6	1.18	9.96	9.22	91.6	
L4671-01	CAR-CON	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
L4671-03	1A	8	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L4671-05	1B	9	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L4671-07	2A	10	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L4671-09	2B	11	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L4672-01	DRUM-1	20	1.00	2.00	2.00	100.0	OIL SAMPLE
L4672-04	DRUM-2	21	1.00	2.00	2.00	100.0	OILY-DEBRIS
L4681-01	TP-1-S1	12	1.14	9.65	8.49	86.4	
L4681-02	TP-1-S2	13	1.13	9.72	8.56	86.5	
L4681-03	TP-1-S	14	1.16	9.86	8.66	86.2	
L4681-05	TP-1-D1	15	1.1	9.89	8.98	89.6	
L4681-06	TP-1-D2	16	1.13	9.82	8.94	89.9	
L4681-07	TP-1-D	17	1.16	9.91	9.22	92.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$