


LB111892

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54764



Linear

A= 0.0000e+000

B= 1.8184e-004

C= 3.2087e-002

Rho= 0.9998900

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	65	0.000	65				
0.2	0.200	0.233	0.033	1105	0.0 %	1105				
2.5	2.500	2.398	-0.102	13009	0.0 %	13009				
5.0	5.000	4.974	-0.026	27178	0.0 %	27178				
7.5	7.500	7.561	0.061	41404	0.0 %	41404				
10.0	10.000	9.990	-0.010	54764	0.0 %	54764				

LB111892 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Nov 2020 14:59:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	65	-	0.0000	7471B	PPB	09 Nov 2020 15:00:56	S	Std
0.2 - 1	50.2	1105	-	0.2000	7471B	PPB	09 Nov 2020 15:03:13	S	Std
2.5 - 1	32.5	13009	-	2.5000	7471B	PPB	09 Nov 2020 15:05:30	S	Std
5.0 - 1	55	27178	-	5.0000	7471B	PPB	09 Nov 2020 15:07:46	S	Std
7.5 - 1	57.5	41404	-	7.5000	7471B	PPB	09 Nov 2020 15:10:03	S	Std
10.0 - 1	170	54764	-	10.0000	7471B	PPB	09 Nov 2020 15:12:19	S	Std
ICV95 - 1	ICV95	20699	3.7960	-	7471B	PPB	09 Nov 2020 15:15:06	U	SMPL
ICB95 - 1	ICB95	-47	0.0235	-	7471B	PPB	09 Nov 2020 15:17:21	U	SMPL
CCV01 - 1	CCV01	27318	4.9996	-	7471B	PPB	09 Nov 2020 15:19:37	U	SMPL
CCB01 - 1	CCB01	-61	0.0210	-	7471B	PPB	09 Nov 2020 15:21:51	U	SMPL
CRA - 1	CRA	1087	0.2297	-	7471B	PPB	09 Nov 2020 15:24:07	U	SMPL
HighStd - 1	HighStd	54959	10.0258	-	7471B	PPB	09 Nov 2020 15:26:22	U	SMPL
ChkStd - 1	ChkStd	39879	7.2837	-	7471B	PPB	09 Nov 2020 15:31:15	U	SMPL
PB132899BL - 1	PBS	-76	0.0183	-	7471B	PPB	09 Nov 2020 15:33:30	U	SMPL
PB132899BS - 1	LCSS	18002	3.3056	-	7471B	PPB	09 Nov 2020 15:35:48	U	SMPL
L4589-01 - 1	CL-02-110620	2602	0.5052	-	7471B	PPB	09 Nov 2020 15:42:53	U	SMPL
L4589-01DUP - 1	CL-02-110620DUP	2487	0.4843	-	7471B	PPB	09 Nov 2020 15:45:09	U	SMPL
L4589-01MS - 1	CL-02-110620MS	20505	3.7607	-	7471B	PPB	09 Nov 2020 15:47:26	U	SMPL
L4589-01MSD - 1	CL-02-110620MSD	20240	3.7125	-	7471B	PPB	09 Nov 2020 15:49:43	U	SMPL
L4658-02 - 1	361-OS	387	0.1025	-	7471B	PPB	09 Nov 2020 15:51:58	U	SMPL
CCV02 - 1	CCV02	26759	4.8979	-	7471B	PPB	09 Nov 2020 15:54:13	U	SMPL
CCB02 - 1	CCB02	-65	0.0203	-	7471B	PPB	09 Nov 2020 15:56:29	U	SMPL
L4658-04 - 1	361-CL	511	0.1250	-	7471B	PPB	09 Nov 2020 15:58:44	U	SMPL
L4658-06 - 1	1998-OS	403	0.1054	-	7471B	PPB	09 Nov 2020 16:00:59	U	SMPL
L4658-08 - 1	1998-CL	965	0.2076	-	7471B	PPB	09 Nov 2020 16:03:15	U	SMPL
L4665-01 - 1	TP-8-COMP	1012	0.2161	-	7471B	PPB	09 Nov 2020 16:05:30	U	SMPL
L4669-01 - 1	TR-04-110620	876	0.1914	-	7471B	PPB	09 Nov 2020 16:07:46	U	SMPL
L4589-01LX5 - 1		477	0.1188	-	7471B	PPB	09 Nov 2020 16:10:02	U	SMPL
L4589-01A - 1		23453	4.2968	-	7471B	PPB	09 Nov 2020 16:12:19	U	SMPL
CCV03 - 1	CCV03	27073	4.9550	-	7471B	PPB	09 Nov 2020 16:14:35	U	SMPL
CCB03 - 1	CCB03	-64	0.0204	-	7471B	PPB	09 Nov 2020 16:16:52	U	SMPL