



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111894

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/09/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	11/09/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	11/09/2020	17:18
4	ICV	1	Water	NA	NA	20.1	7.00	11/09/2020	17:19
5	CCV1	1	Water	NA	NA	20.2	2.02	11/09/2020	17:20
6	L4672-02	1	Solid	20.03	20	20.1	5.55	11/09/2020	17:21
7	L4672-05	1	Solid	20.03	20	19.6	5.56	11/09/2020	17:22
8	L4680-02	1	Solid	20.04	20	19.4	9.22	11/09/2020	17:24
9	L4680-04	1	Solid	20.03	20	19.2	9.29	11/09/2020	17:25
10	L4680-06	1	Solid	20.03	20	19.2	8.66	11/09/2020	17:26
11	L4681-04	1	Solid	20.04	20	21.4	6.08	11/09/2020	17:29
12	L4681-08	1	Solid	20.03	20	21.5	6.43	11/09/2020	17:30
13	L4681-08DUP	1	Solid	20.04	20	21.6	6.44	11/09/2020	17:31
14	CCV2	1	Water	NA	NA	20.2	12.02	11/09/2020	17:33

13 11894

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4681 WorkList ID : 144014 Department : Wet-Chemistry Date : 11-09-2020 15:43:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2020	Solid	L4672-02	Corrosivity	Cool 4 deg C	PSEG03	F12	DRUM-1	11/09/2020	9045D
11/09/2020	Solid	L4672-05	Corrosivity	Cool 4 deg C	PSEG03	F12	DRUM-2	11/09/2020	9045D
11/05/2020	Solid	L4680-02	Corrosivity	Cool 4 deg C	ENTE01	F51	PAV-B7-110520-0-10	11/05/2020	9045D
11/05/2020	Solid	L4680-04	Corrosivity	Cool 4 deg C	ENTE01	F51	PAV-B7-110520-DP	11/05/2020	9045D
11/05/2020	Solid	L4680-06	Corrosivity	Cool 4 deg C	ENTE01	F51	PAV-B7-110920-10-20	11/05/2020	9045D
11/09/2020	Solid	L4681-04	Corrosivity	Cool 4 deg C	PSEG03	F12	TP-1-S	11/09/2020	9045D
11/09/2020	Solid	L4681-08	Corrosivity	Cool 4 deg C	PSEG03	F12	TP-1-D	11/09/2020	9045D

Date/Time 11-09-2020 16:46
 Received by: 70 (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 11-09-2020 18:20
 Received by: KS (wet-chem)
 Relinquished by: 70 (wet-chem)