


LB 111895

ISM02.4

INSTRUMENT ID: CV1

 μ Abs.:

39524



Linear

A= 0.0000e+000

B= 2.5562e-004

C= -1.0455e-002

Rho= 0.9997508

Accept=Accepted

Slope
y=mx+cepf

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>σ/μ</i>
0.00	0.000	0.004	0.004	56	0.000	56					<i>1</i>
0.05	0.050					244					<i>1</i>
0.20	0.200	0.199	-0.001	821	0.0 %	821					<i>1</i>
2.50	2.500	2.525	0.025	9917	0.0 %	9917					<i>1</i>
5.00	5.000	5.050	0.050	19797	0.0 %	19797					<i>2</i>
7.50	7.500	7.330	-0.170	28715	0.0 %	28715					<i>1</i>
10.0	10.000	10.093	0.093	39524	0.0 %	39524					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	56	-		0 ISM02.4	PPB	11/10/2020 11:27	Std
	0.2 S0.2	821	-		0.2 ISM02.4	PPB	11/10/2020 11:29	Std
	2.5 S2.5	9917	-		2.5 ISM02.4	PPB	11/10/2020 11:31	Std
	5 S5.0	19797	-		5 ISM02.4	PPB	11/10/2020 11:33	Std
	7.5 S7.5	28715	-		7.5 ISM02.4	PPB	11/10/2020 11:36	Std
	10 S10.0	39524	-		10 ISM02.4	PPB	11/10/2020 11:40	Std
ICV	ICV	15131	3.8573 -		ISM02.4	PPB	11/10/2020 11:43	SMPL
ICB	ICB	-83	-0.0317 -		ISM02.4	PPB	11/10/2020 11:45	SMPL
CCV004	CCV004	19634	5.0083 -		ISM02.4	PPB	11/10/2020 11:48	SMPL
CCB004	CCB004	-72	-0.0289 -		ISM02.4	PPB	11/10/2020 11:50	SMPL
PB132903BL	PBW903	-16	-0.0145 -		ISM02.4	PPB	11/10/2020 11:52	SMPL
PB132903TB	LEB903	24	-0.0043 -		ISM02.4	PPB	11/10/2020 11:54	SMPL
L4663-01	MBG0R1	178	0.035 -		ISM02.4	PPB	11/10/2020 11:57	SMPL
L4663-02	MBG0R1D	148	0.0274 -		ISM02.4	PPB	11/10/2020 11:59	SMPL
L4663-03	MBG0R1S	3900	0.9865 -		ISM02.4	PPB	11/10/2020 12:01	SMPL
CCV005	CCV005	19363	4.9391 -		ISM02.4	PPB	11/10/2020 12:06	SMPL
CCB005	CCB005	-65	-0.0271 -		ISM02.4	PPB	11/10/2020 12:08	SMPL