


LB111898

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4931e-004

C= 5.2386e-003

Rho= 0.9996051

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/01
0.00	0.000	0.011	0.011	23	0.000	23					-
0.05	0.050					244					-
0.20	0.200	0.234	0.034	918	0.0 %	918					17
2.50	2.500	2.559	0.059	10244	0.0 %	10244					2
5.00	5.000	4.924	-0.076	19731	0.0 %	19731					-2
7.50	7.500	7.322	-0.178	29348	0.0 %	29348					-2
10.0	10.000	10.149	0.149	40689	0.0 %	40689					1

LB111898
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	23	-		0 ISM02.4	PPB	11/10/2020 13:37	Std
	0.2 S0.2	918	-		0.2 ISM02.4	PPB	11/10/2020 13:39	Std
	2.5 S2.5	10244	-		2.5 ISM02.4	PPB	11/10/2020 13:41	Std
	5 S5.0	19731	-		5 ISM02.4	PPB	11/10/2020 13:43	Std
	7.5 S7.5	29348	-		7.5 ISM02.4	PPB	11/10/2020 13:46	Std
	10 S10.0	40689	-		10 ISM02.4	PPB	11/10/2020 13:48	Std
ICV	ICV	14869	3.7122 -		ISM02.4	PPB	11/10/2020 13:51	SMPL
ICB	ICB	-78	-0.0142 -		ISM02.4	PPB	11/10/2020 13:53	SMPL
CCV006	CCV006	19655	4.9054 -		ISM02.4	PPB	11/10/2020 13:55	SMPL
CCB006	CCB006	-89	-0.0169 -		ISM02.4	PPB	11/10/2020 13:57	SMPL
PB132927BL	PBW927	-8	0.0032 -		ISM02.4	PPB	11/10/2020 14:00	SMPL
L4664-01	MBG0R2	-26	-0.0012 -		ISM02.4	PPB	11/10/2020 14:02	SMPL
L4664-02	MBG0R2D	1	0.0055 -		ISM02.4	PPB	11/10/2020 14:04	SMPL
L4664-03	MBG0R2S	3831	0.9603 -		ISM02.4	PPB	11/10/2020 14:06	SMPL
CCV007	CCV007	18692	4.6653 -		ISM02.4	PPB	11/10/2020 14:09	SMPL
CCB007	CCB007	-74	-0.0132 -		ISM02.4	PPB	11/10/2020 14:11	SMPL