

LB111915


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

31701



A= 0.0000e+000

B= 3.1913e-004

C= -1.3893e-002

Rho= 0.9995239

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	36	0.000	36				
0.2	0.200	0.177	-0.023	597	0.0 %	597				
2.5	2.500	2.555	0.055	8049	0.0 %	8049				
5.0	5.000	5.100	0.100	16024	0.0 %	16024				
7.5	7.500	7.268	-0.232	22819	0.0 %	22819				
10.0	10.000	10.103	0.103	31701	0.0 %	31701				

LB111915 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Nov 2020 10:50:37

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.0000	7470A	PPB	11 Nov 2020 10:56:29	S	Std
0.2 - 1	50.2	597	-	0.2000	7470A	PPB	11 Nov 2020 10:58:47	S	Std
2.5 - 1	52.5	8049	-	2.5000	7470A	PPB	11 Nov 2020 11:01:04	S	Std
5.0 - 1	55	16024	-	5.0000	7470A	PPB	11 Nov 2020 11:03:21	S	Std
7.5 - 1	57.5	22819	-	7.5000	7470A	PPB	11 Nov 2020 11:05:38	S	Std
10.0 - 1	510	31701	-	10.0000	7470A	PPB	11 Nov 2020 11:07:54	S	Std
ICV97 - 1	ICV97	11352	3.6089	-	7470A	PPB	11 Nov 2020 11:10:35	U	SMPL
ICB97 - 1	ICB97	-69	-0.0359	-	7470A	PPB	11 Nov 2020 11:12:50	U	SMPL
CCV11 - 1	CCV11	15606	4.9664	-	7470A	PPB	11 Nov 2020 11:15:05	U	SMPL
CCB11 - 1	CCB11	-68	-0.0356	-	7470A	PPB	11 Nov 2020 11:17:19	U	SMPL
CRA - 1	CRA	563	0.1658	-	7470A	PPB	11 Nov 2020 11:19:34	U	SMPL
HighStd - 1	HighStd	33813	10.7768	-	7470A	PPB	11 Nov 2020 11:21:49	U	SMPL
ChkStd - 1	ChkStd	21614	6.8838	-	7470A	PPB	11 Nov 2020 11:24:04	U	SMPL
PB132929BL - 1	PBW	-132	-0.0560	-	7470A	PPB	11 Nov 2020 11:26:21	U	SMPL
PB132929BS - 1	LCSW	11880	3.7774	-	7470A	PPB	11 Nov 2020 11:28:37	U	SMPL
L4680-02 - 1	PAV-B7-110520-0-10	-71	-0.0366	-	7470A	PPB	11 Nov 2020 11:30:53	U	SMPL
L4680-04 - 1	PAV-B7-110520-DP	-4	-0.0152	-	7470A	PPB	11 Nov 2020 11:33:10	U	SMPL
L4680-06 - 1	PAV-B7-110920-10-20	21	-0.0072	-	7470A	PPB	11 Nov 2020 11:35:27	U	SMPL
L4681-04 - 1	TP-1-S	13	-0.0097	-	7470A	PPB	11 Nov 2020 11:37:43	U	SMPL
L4681-08 - 1	TP-1-D	7	-0.0117	-	7470A	PPB	11 Nov 2020 11:39:59	U	SMPL
CCV12 - 1	CCV12	16435	5.2310	-	7470A	PPB	11 Nov 2020 11:42:14	U	SMPL
CCB12 - 1	CCB12	-50	-0.0298	-	7470A	PPB	11 Nov 2020 11:44:29	U	SMPL
L4681-08DUP - 1	TP-1-DDUP	0	-0.0139	-	7470A	PPB	11 Nov 2020 11:46:44	U	SMPL
L4681-08MS - 1	TP-1-DMS	13269	4.2206	-	7470A	PPB	11 Nov 2020 11:48:59	U	SMPL
L4681-08MSD - 1	TP-1-DMSD	11851	3.7681	-	7470A	PPB	11 Nov 2020 11:51:15	U	SMPL
PB132929TB - 1	PB132904TB	-53	-0.0308	-	7470A	PPB	11 Nov 2020 11:53:30	U	SMPL
L4681-08LX5 - 1		-1	-0.0142	-	7470A	PPB	11 Nov 2020 11:55:47	U	SMPL
L4681-08A - 1		13542	4.3078	-	7470A	PPB	11 Nov 2020 11:58:03	U	SMPL
CCV13 - 1	CCV13	16542	5.2651	-	7470A	PPB	11 Nov 2020 12:00:20	U	SMPL
CCB13 - 1	CCB13	-55	-0.0314	-	7470A	PPB	11 Nov 2020 12:02:37	U	SMPL