



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111918

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP88120
Turbidity Calibration std, 20NTU	WP88116
Turbidity Calibration std, 40NTU	WP88115
Turbidity Calibration std, 1NTU	WP88119
Turbidity Calibration std, 5NTU	WP88118
Turbidity Calibration - CCV std, 10 NTU	WP88117
10 NTU Standard 500 ml	W2752

Intercept: 0.2706

Slope: 0.9721

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.264	11/11/2020	10:40
2	CAL2	1	1	1.170	11/11/2020	10:42
3	CAL3	5	1	5.320	11/11/2020	10:44
4	CAL4	10	1	9.950	11/11/2020	10:46
5	CAL5	20	1	19.600	11/11/2020	10:48
6	CAL6	40	1	39.200	11/11/2020	10:50



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB111918

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.420	11/11/2020	10:52
2	ICB		1	0.282	0.012	11/11/2020	10:55
3	CCV1	10	1	9.870	9.875	11/11/2020	10:58
4	CCB1		1	0.279	0.009	11/11/2020	11:00
5	LB111918BL		1	0.273	0.002	11/11/2020	11:02
6	L4703-01		1	26.200	26.674	11/11/2020	11:05
7	L4703-01DUP		1	28.500	29.040	11/11/2020	11:08
8	CCV2	10	1	9.760	9.762	11/11/2020	11:11
9	CCB2		1	0.275	0.005	11/11/2020	11:14

LB111918
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-111120 WorkList ID : 144068 Department : Wet-Chemistry Date : 11-11-2020 10:01:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2020	Water	L4703-01	Turbidity	Cool 4 deg C	LIRO01	F51	UDS-4-40	11/10/2020	SM2130 B

Date/Time 11/11/2020 10:15
 Received by: AMC wet-chem
 Relinquished by: JH Coelt - Chem

Date/Time 11/11/2020 11:20
 Received by: JH Coelt - Chem
 Relinquished by: AMC wet-chem