



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111919

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/11/2020	10:45
2	CAL2	1	Water	NA	NA	20.1	7.00	11/11/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	11/11/2020	10:42
4	ICV	1	Water	NA	NA	20.2	6.99	11/11/2020	10:43
5	CCV1	1	Water	NA	NA	20.2	2.01	11/11/2020	10:45
6	L4683-01	1	Solid	20.03	20	21.9	8.43	11/11/2020	10:46
7	L4683-05	1	Solid	20.04	20	21.8	2.85	11/11/2020	10:48
8	L4683-06	1	Solid	20.01	20	22.8	7.69	11/11/2020	10:49
9	L4702-05	1	Solid	20.03	20	22.4	7.08	11/11/2020	11:00
10	L4702-06	1	Solid	20.04	20	22.4	8.69	11/11/2020	11:02
11	L4702-07	1	Solid	20.03	20	22.3	9.39	11/11/2020	11:03
12	L4702-08	1	Solid	20.04	20	22.4	9.37	11/11/2020	11:05
13	L4702-08DUP	1	Solid	20.03	20	22.5	9.38	11/11/2020	11:06
14	CCV2	1	Water	NA	NA	20.2	12.02	11/11/2020	11:09

VB111919

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I3686 WorkList ID : 144072 Department : Wet-Chemistry Date : 11-11-2020 10:35:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/10/2020	Solid	L4683-01	Corrosivity	Cool 4 deg C	PSEG03	F22	OILY-DEBRIS-01-02-03	11/10/2020	9045D
11/10/2020	Solid	L4683-05	Corrosivity	Cool 4 deg C	PSEG03	F22	TOTE	11/10/2020	9045D
11/10/2020	Solid	L4683-06	Corrosivity	Cool 4 deg C	PSEG03	F22	OILY-WATER-001-002-003	11/10/2020	9045D
11/10/2020	Solid	L4702-05	Corrosivity	Cool 4 deg C	WSP001	F61	M-5(0-10)	11/10/2020	9045D
11/10/2020	Solid	L4702-06	Corrosivity	Cool 4 deg C	WSP001	F61	M-5(10-20)	11/10/2020	9045D
11/10/2020	Solid	L4702-07	Corrosivity	Cool 4 deg C	WSP001	F61	M-6(0-10)	11/10/2020	9045D
11/10/2020	Solid	L4702-08	Corrosivity	Cool 4 deg C	WSP001	F61	M-99(0-10)	11/10/2020	9045D

Date/Time 11-11-2020 10:40
 Received by: Jo (wet Chem)
 Relinquished by: Jo (wet Chem)

Date/Time 11-11-2020 15:00
 Received by: Jo (wet Chem)
 Relinquished by: Jo (wet Chem)