



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111922

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/11/2020	10:45
2	CAL2	1	Water	NA	NA	20.1	7.00	11/11/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	11/11/2020	10:42
4	ICV	1	Water	NA	NA	20.2	6.99	11/11/2020	10:43
5	CCV1	1	Water	NA	NA	20.2	2.01	11/11/2020	10:45
6	L4699-01	1	Solid	20.03	20	22.1	7.54	11/11/2020	11:10
7	L4699-01DUP	1	Solid	20.04	20	22.2	7.55	11/11/2020	11:11
8	CCV2	1	Water	NA	NA	20.2	12.02	11/11/2020	11:12

1B111922

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14599

WorkList ID : 144073

Department : Wet-Chemistry

Date : 11-11-2020 11:53:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/10/2020	Solid	L4699-01-1	pH	Cool 4 deg C	TULL01	F51	STOCKPILE	11/10/2020	9045D

Date/Time 11.11.2020 12:00
Received by: SB (wet Chem)
Relinquished by: SB

Date/Time 11.11.2020 13:00
Received by: SB
Relinquished by: SB (wet Chem)