


LB111935

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	8	0.000	8				
0.2	0.200	0.206	0.006	643	0.0 %	643				
2.5	2.500	2.452	-0.048	7936	0.0 %	7936				
5.0	5.000	4.930	-0.070	15980	0.0 %	15980				
7.5	7.500	7.757	0.257	25155	0.0 %	25155				
10.0	10.000	9.844	-0.156	31933	0.0 %	31933				

LB111935 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Nov 2020 09:43:56

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	8	-	0.0000	7470A	PPB	12 Nov 2020 09:56:17	S	Std
0.2 - 1	SD 2	643	-	0.2000	7470A	PPB	12 Nov 2020 09:58:34	S	Std
2.5 - 1	SD 5	7936	-	2.5000	7470A	PPB	12 Nov 2020 10:00:51	S	Std
5.0 - 1	SD 5	15980	-	5.0000	7470A	PPB	12 Nov 2020 10:03:08	S	Std
7.5 - 1	SD 5	25155	-	7.5000	7470A	PPB	12 Nov 2020 10:05:24	S	Std
10.0 - 1	SD 10	31933	-	10.0000	7470A	PPB	12 Nov 2020 10:10:09	S	Std
ICV99 - 1	ICV99	13113	4.0472	-7470A	PPB	12 Nov 2020 10:12:49	U	SMPL	
ICB99 - 1	ICB99	-79	-0.0165	-7470A	PPB	12 Nov 2020 10:15:04	U	SMPL	
CCV17 - 1	CCV17	16353	5.0452	-7470A	PPB	12 Nov 2020 10:17:19	U	SMPL	
CCB17 - 1	CCB17	-90	-0.0199	-7470A	PPB	12 Nov 2020 10:19:34	U	SMPL	
CRA - 1	CRA	633	0.2028	-7470A	PPB	12 Nov 2020 10:21:49	U	SMPL	
HighStd - 1	HighStd	31850	9.8189	-7470A	PPB	12 Nov 2020 10:24:04	U	SMPL	
ChkStd - 1	ChkStd	23104	7.1248	-7470A	PPB	12 Nov 2020 10:26:19	U	SMPL	
PB132950BL - 1	PBW	-125	-0.0307	-7470A	PPB	12 Nov 2020 10:28:35	U	SMPL	
PB132950BS - 1	LCSW	14680	4.5299	-7470A	PPB	12 Nov 2020 10:30:52	U	SMPL	
L4622-04 - 1	PAV-B12-0-10	-81	-0.0171	-7470A	PPB	12 Nov 2020 10:33:08	U	SMPL	
L4622-06 - 1	PAV-B12-10-17	5	0.0094	-7470A	PPB	12 Nov 2020 10:35:24	U	SMPL	
L4631-04 - 1	PAV-B11-110420-0-10	9	0.0106	-7470A	PPB	12 Nov 2020 10:37:41	U	SMPL	
L4631-06 - 1	PAV-B11-110420-10-15.25	2	0.0085	-7470A	PPB	12 Nov 2020 10:39:58	U	SMPL	
L4697-04 - 1	TP-2-S	23	0.0149	-7470A	PPB	12 Nov 2020 10:42:13	U	SMPL	
CCV18 - 1	CCV18	16077	4.9602	-7470A	PPB	12 Nov 2020 10:44:28	U	SMPL	
CCB18 - 1	CCB18	-57	-0.0097	-7470A	PPB	12 Nov 2020 10:46:44	U	SMPL	
L4697-08 - 1	TP-2-S	-6	0.0060	-7470A	PPB	12 Nov 2020 10:48:59	U	SMPL	
L4701-02 - 1	TRANSFORMER-1-(1-4)	134	0.0491	-7470A	PPB	12 Nov 2020 10:51:14	U	SMPL	
L4701-13 - 1	TRANSFORMER-5-(1-4)	18	0.0134	-7470A	PPB	12 Nov 2020 10:53:29	U	SMPL	
L4701-13DUP - 1	TRANSFORMER-5-(1-4)DUP	33	0.0180	-7470A	PPB	12 Nov 2020 10:55:45	U	SMPL	
L4701-13MS - 1	TRANSFORMER-5-(1-4)MS	11719	3.6178	-7470A	PPB	12 Nov 2020 10:58:00	U	SMPL	
L4701-13MSD - 1	TRANSFORMER-5-(1-4)MSD	11189	3.4545	-7470A	PPB	12 Nov 2020 11:00:17	U	SMPL	
PB132950TB - 1	PB132933TB	-44	-0.0057	-7470A	PPB	12 Nov 2020 11:02:33	U	SMPL	
L4701-13LX5 - 1		7	0.0100	-7470A	PPB	12 Nov 2020 11:04:50	U	SMPL	
L4701-13A - 1		12625	3.8968	-7470A	PPB	12 Nov 2020 11:07:07	U	SMPL	
CCV19 - 1	CCV19	15902	4.9063	-7470A	PPB	12 Nov 2020 11:09:22	U	SMPL	
CCB19 - 1	CCB19	-65	-0.0122	-7470A	PPB	12 Nov 2020 11:11:38	U	SMPL	