



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111950

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	11/12/2020	13:45
2	L4683-03		1	8	100.00	>212.0	>212.0	11/12/2020	14:20
3	L4683-03DUP		1	8	100.00	>212.0	>212.0	11/12/2020	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB111950

WorkList Name : fp111220 WorkList ID : 144101 Department : Wet-Chemistry Date : 11-12-2020 08:58:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2020	Water	L4683-03	Flash Point	Cool 4 deg C	PSEG03	F22	TOTE	11/10/2020	1010A

Date/Time 11/12/2020 13:15
 Received by: KS (Wetchem)
 Relinquished by: JR (Wetchem)

Date/Time 11/12/2020 14:55
 Received by: JR (Wetchem)
 Relinquished by: KS (Wetchem)