


LB111955

7470A

INSTRUMENT ID: CV1

μ Abs.:

36145



Linear

A= 0.0000e+000

B= 2.7690e-004

C= -2.7119e-003

Rho= 0.9999897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	20	0.000	20				
0.2	0.200	0.184	-0.016	674	0.0 %	674				
2.5	2.500	2.532	0.032	9153	0.0 %	9153				
5.0	5.000	4.981	-0.019	18000	0.0 %	18000				
7.5	7.500	7.494	-0.006	27075	0.0 %	27075				
10.0	10.000	10.006	0.006	36145	0.0 %	36145				

LB111955 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Nov 2020 09:06:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7470A	PPB	13 Nov 2020 09:09:07	S	Std
0.2 - 1	50.2	674	-	0.2000	7470A	PPB	13 Nov 2020 09:11:24	S	Std
2.5 - 1	52.5	9153	-	2.5000	7470A	PPB	13 Nov 2020 09:13:41	S	Std
5.0 - 1	55	18000	-	5.0000	7470A	PPB	13 Nov 2020 09:15:57	S	Std
7.5 - 1	57.5	27075	-	7.5000	7470A	PPB	13 Nov 2020 09:18:14	S	Std
10.0 - 1	510	36145	-	10.0000	7470A	PPB	13 Nov 2020 09:20:29	S	Std
ICV02 - 1	ICV02	13923	3.8525	-	7470A	PPB	13 Nov 2020 09:23:10	U	SMPL
ICB02 - 1	ICB02	-77	-0.0240	-	7470A	PPB	13 Nov 2020 09:25:25	U	SMPL
CCV24 - 1	CCV24	18317	5.0692	-	7470A	PPB	13 Nov 2020 09:27:40	U	SMPL
CCB24 - 1	CCB24	-87	-0.0268	-	7470A	PPB	13 Nov 2020 09:29:55	U	SMPL
CRA - 1	CRA	685	0.1870	-	7470A	PPB	13 Nov 2020 09:32:10	U	SMPL
HighStd - 1	HighStd	36586	10.1279	-	7470A	PPB	13 Nov 2020 09:34:24	U	SMPL
ChkStd - 1	ChkStd	24226	6.7054	-	7470A	PPB	13 Nov 2020 09:36:39	U	SMPL
PB132977BL - 1	PBW	-120	-0.0359	-	7470A	PPB	13 Nov 2020 09:38:57	U	SMPL
PB132977BS - 1	LCSW	12856	3.5571	-	7470A	PPB	13 Nov 2020 09:41:12	U	SMPL
L4699-02 - 1	STOCKPILE	-56	-0.0182	-	7470A	PPB	13 Nov 2020 09:43:28	U	SMPL
L4702-05 - 1	M-5(0-10)	8	-0.0005	-	7470A	PPB	13 Nov 2020 09:45:45	U	SMPL
L4702-06 - 1	M-5(10-20)	10	0.0001	-	7470A	PPB	13 Nov 2020 09:48:02	U	SMPL
L4702-07 - 1	M-6(0-10)	19	0.0025	-	7470A	PPB	13 Nov 2020 09:50:18	U	SMPL
L4702-08 - 1	M-99(0-10)	61	0.0142	-	7470A	PPB	13 Nov 2020 09:52:34	U	SMPL
CCV25 - 1	CCV25	18005	4.9828	-	7470A	PPB	13 Nov 2020 09:54:49	U	SMPL
CCB25 - 1	CCB25	-58	-0.0188	-	7470A	PPB	13 Nov 2020 09:57:04	U	SMPL
L4705-01 - 1	COMP-1	15	0.0014	-	7470A	PPB	13 Nov 2020 09:59:20	U	SMPL
L4705-03 - 1	COMP-2	25	0.0042	-	7470A	PPB	13 Nov 2020 10:01:35	U	SMPL
L4705-05 - 1	32785	8	-0.0005	-	7470A	PPB	13 Nov 2020 10:03:50	U	SMPL
L4715-02 - 1	M-6(10-20)	43	0.0092	-	7470A	PPB	13 Nov 2020 10:06:06	U	SMPL
L4719-04 - 1	TP-3-S	1	-0.0024	-	7470A	PPB	13 Nov 2020 10:08:21	U	SMPL
L4719-08 - 1	TP-3-D	-3	-0.0035	-	7470A	PPB	13 Nov 2020 10:10:38	U	SMPL
L4719-12 - 1	TP-3A-S	17	0.0020	-	7470A	PPB	13 Nov 2020 10:12:54	U	SMPL
L4719-12DUP - 1	TP-3A-SDUP	8	-0.0005	-	7470A	PPB	13 Nov 2020 10:15:11	U	SMPL
L4719-12MS - 1	TP-3A-SMS	12816	3.5460	-	7470A	PPB	13 Nov 2020 10:17:28	U	SMPL
L4719-12MSD - 1	TP-3A-SMSD	13689	3.7877	-	7470A	PPB	13 Nov 2020 10:19:43	U	SMPL
CCV26 - 1	CCV26	18081	5.0039	-	7470A	PPB	13 Nov 2020 10:21:59	U	SMPL
CCB26 - 1	CCB26	-94	-0.0287	-	7470A	PPB	13 Nov 2020 10:24:14	U	SMPL
PB132977TB - 1	PB132946TB	0	-0.0027	-	7470A	PPB	13 Nov 2020 10:26:30	U	SMPL
L4719-12LX5 - 1		2	-0.0022	-	7470A	PPB	13 Nov 2020 10:28:45	U	SMPL
L4719-12A - 1		13624	3.7697	-	7470A	PPB	13 Nov 2020 10:31:01	U	SMPL
CCV27 - 1	CCV27	17951	4.9679	-	7470A	PPB	13 Nov 2020 10:33:16	U	SMPL
CCB27 - 1	CCB27	-62	-0.0199	-	7470A	PPB	13 Nov 2020 10:35:32	U	SMPL