



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111956

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/13/2020	10:22
2	CAL2	1	Water	NA	NA	20.1	7.00	11/13/2020	10:23
3	CAL3	1	Water	NA	NA	20.1	10.02	11/13/2020	10:24
4	ICV	1	Water	NA	NA	20.2	7.01	11/13/2020	10:25
5	CCV1	1	Water	NA	NA	20.1	2.02	11/13/2020	10:26
6	L4736-01	1	Water	NA	NA	19.5	6.20	11/13/2020	10:27
7	L4736-01DUP	1	Water	NA	NA	19.6	6.21	11/13/2020	10:28
8	CCV2	1	Water	NA	NA	20.1	12.02	11/13/2020	10:30

LB 111956

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14736 WorkList ID : 144138 Department : Wet-Chemistry Date : 11-13-2020 10:17:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/10/2020	Water	L4736-01	-pH	Cool 4 deg C	STER06	F51	POOL-WATER	11/10/2020	9040C

Date/Time 11-13-2020 10:20
Received by: JH Cufft-chem
Relinquished by: JH Cufft-chem

Date/Time 11-13-2020 11:00
Received by: JH Cufft-chem
Relinquished by: JH Cufft-chem