



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111960

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/13/2020	09:35
2	L4672-02		1	8	100.00	>212.0	>212.0	11/13/2020	10:05
3	L4672-02DUP		1	8	100.00	>212.0	>212.0	11/13/2020	10:40
4	L4683-06		1	9	100.00	>212.0	>212.0	11/13/2020	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB111960

WorkList Name : fp111220-1 WorkList ID : 144102 Department : Wet-Chemistry Date : 11-12-2020 09:00:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2020	Solid	L4672-02	Flash Point	Cool 4 deg C	PSEG03	F12	DRUM-1	11/09/2020	1010A
11/17/2020	Solid	L4683-06	Flash Point	Cool 4 deg C	PSEG03	F22	OILY-WATER-001-002-003	11/10/2020	1010A

Date/Time 11/13/2020 09:05
 Received by: KS (wetcher)
 Relinquished by: T. Claf Allen

Date/Time 11/13/2020 11:25
 Received by: T. Claf Allen
 Relinquished by: KS (wetcher)