


LB111962

INSTRUMENT ID: CV1

245.1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	123	0.000	123				
0.05	0.050	0.057	0.007	301	0.0 %	301	0			
0.2	0.200	0.203	0.003	805	0.0 %	805				
2.5	2.500	2.440	-0.060	8531	0.0 %	8531				
5.0	5.000	5.045	0.045	17529	0.0 %	17529				
7.5	7.500	7.537	0.037	26137	0.0 %	26137				
10.0	10.000	9.964	-0.036	34517	0.0 %	34517				

LB111962 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Nov 2020 13:35:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	123	-	0.0000	245.1	PPB	13 Nov 2020 13:35:45	S	Std
0.05 - 1	50.05	301	-	0.0500	245.1	PPB	13 Nov 2020 13:38:03	S	Std
0.2 - 1	50.2	805	-	0.2000	245.1	PPB	13 Nov 2020 13:40:20	S	Std
2.5 - 1	52.5	8531	-	2.5000	245.1	PPB	13 Nov 2020 13:42:37	S	Std
5.0 - 1	55	17529	-	5.0000	245.1	PPB	13 Nov 2020 13:44:53	S	Std
7.5 - 1	57.5	26137	-	7.5000	245.1	PPB	13 Nov 2020 13:47:08	S	Std
10.0 - 1	510	34517	-	10.0000	245.1	PPB	13 Nov 2020 13:51:58	S	Std
ICV03 - 1	ICV03	13270	3.8118	-245.1		PPB	13 Nov 2020 13:54:39	U	SMPL
ICB03 - 1	ICB03	58	-0.0137	-245.1		PPB	13 Nov 2020 13:59:26	U	SMPL
CCV28 - 1	CCV28	17706	5.0962	-245.1		PPB	13 Nov 2020 14:01:41	U	SMPL
CCB28 - 1	CCB28	-47	-0.0441	-245.1		PPB	13 Nov 2020 14:03:56	U	SMPL
CRA - 1	CRA	260	0.0448	-245.1		PPB	13 Nov 2020 14:06:12	U	SMPL
HighStd - 1	HighStd	34769	10.0367	-245.1		PPB	13 Nov 2020 14:08:27	U	SMPL
ChkStd - 1	ChkStd	24468	7.0541	-245.1		PPB	13 Nov 2020 14:10:42	U	SMPL
PB132982BL - 1	PBW	-23	-0.0371	-245.1		PPB	13 Nov 2020 14:12:59	U	SMPL
PB132982BS - 1	LCSW	12039	3.4553	-245.1		PPB	13 Nov 2020 14:15:16	U	SMPL
L4632-02 - 1	COMP	568	0.1340	-245.1		PPB	13 Nov 2020 14:17:32	U	SMPL
L4632-02DUP - 1	COMPDUP	522	0.1207	-245.1		PPB	13 Nov 2020 14:19:48	U	SMPL
L4632-02MS - 1	COMPMS	11889	3.4119	-245.1		PPB	13 Nov 2020 14:25:17	U	SMPL
L4632-02MSD - 1	COMPMSD	12429	3.5683	-245.1		PPB	13 Nov 2020 14:27:33	U	SMPL
L4728-03 - 1	NCQ-632	66	-0.0114	-245.1		PPB	13 Nov 2020 14:29:48	U	SMPL
CCV29 - 1	CCV29	17722	5.1008	-245.1		PPB	13 Nov 2020 14:32:04	U	SMPL
CCB29 - 1	CCB29	-64	-0.0490	-245.1		PPB	13 Nov 2020 14:34:19	U	SMPL
L4729-03 - 1	TI-658	133	0.0080	-245.1		PPB	13 Nov 2020 14:36:35	U	SMPL
L4730-03 - 1	TI-604	268	0.0471	-245.1		PPB	13 Nov 2020 14:38:51	U	SMPL
L4731-03 - 1	TI-633	242	0.0396	-245.1		PPB	13 Nov 2020 14:41:06	U	SMPL
L4632-02LX5 - 1		-9	-0.0331	-245.1		PPB	13 Nov 2020 14:43:22	U	SMPL
L4632-02A - 1		12394	3.5581	-245.1		PPB	13 Nov 2020 14:48:13	U	SMPL
CCV30 - 1	CCV30	17437	5.0183	-245.1		PPB	13 Nov 2020 14:50:29	U	SMPL
CCB30 - 1	CCB30	-65	-0.0493	-245.1		PPB	13 Nov 2020 14:52:45	U	SMPL