

1311986


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

36880



A= 0.0000e+000

B= 2.7396e-004

C= -3.5303e-002

Rho= 0.9995956

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	43	0.000	43				
0.2	0.200	0.181	-0.019	791	0.0 %	791				
2.5	2.500	2.543	0.043	9413	0.0 %	9413				
5.0	5.000	5.133	0.133	18864	0.0 %	18864				
7.5	7.500	7.298	-0.202	26766	0.0 %	26766				
10.0	10.000	10.068	0.068	36880	0.0 %	36880				

LB111986 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Nov 2020 14:17:52

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	43	-	0.00007470A	PPB	16 Nov 2020 14:21:23	S	Std
0.2 - 1	50.2	791	-	0.20007470A	PPB	16 Nov 2020 14:23:41	S	Std
2.5 - 1	52.5	9413	-	2.50007470A	PPB	16 Nov 2020 14:25:59	S	Std
5.0 - 1	55	18864	-	5.00007470A	PPB	16 Nov 2020 14:28:16	S	Std
7.5 - 1	57.5	26766	-	7.50007470A	PPB	16 Nov 2020 14:30:32	S	Std
10.0 - 1	510	36880	-	10.00007470A	PPB	16 Nov 2020 14:32:50	S	Std
ICV06 - 1	ICV06	14265	3.8728	-7470A	PPB	16 Nov 2020 14:35:38	U	SMPL
ICB06 - 1	ICB06	-81	-0.0575	-7470A	PPB	16 Nov 2020 14:37:53	U	SMPL
CCV40 - 1	CCV40	18643	5.0722	-7470A	PPB	16 Nov 2020 14:40:08	U	SMPL
CCB40 - 1	CCB40	-62	-0.0523	-7470A	PPB	16 Nov 2020 14:42:23	U	SMPL
CRA - 1	CRA	768	0.1751	-7470A	PPB	16 Nov 2020 14:44:38	U	SMPL
HighStd - 1	HighStd	35823	9.7788	-7470A	PPB	16 Nov 2020 14:46:53	U	SMPL
ChkStd - 1	ChkStd	24870	6.7781	-7470A	PPB	16 Nov 2020 14:49:08	U	SMPL
PB133038BL - 1	PBW	-120	-0.0682	-7470A	PPB	16 Nov 2020 14:51:25	U	SMPL
PB133038BS - 1	LCSW	13813	3.7489	-7470A	PPB	16 Nov 2020 14:53:41	U	SMPL
L4734-03 - 1	BARNUM-STREET-TEST-PIT-1	-49	-0.0487	-7470A	PPB	16 Nov 2020 14:55:57	U	SMPL
L4734-04 - 1	BARNUM-STREET-TEST-PIT-2	-5	-0.0367	-7470A	PPB	16 Nov 2020 14:58:14	U	SMPL
L4741-01 - 1	11-13-A	10	-0.0326	-7470A	PPB	16 Nov 2020 15:00:31	U	SMPL
L4741-02 - 1	11-13-B	15	-0.0312	-7470A	PPB	16 Nov 2020 15:02:47	U	SMPL
L4741-03 - 1	11-13-C	20	-0.0298	-7470A	PPB	16 Nov 2020 15:05:03	U	SMPL
CCV41 - 1	CCV41	18685	5.0837	-7470A	PPB	16 Nov 2020 15:07:18	U	SMPL
CCB41 - 1	CCB41	-43	-0.0471	-7470A	PPB	16 Nov 2020 15:09:33	U	SMPL
L4741-03DUP - 1	11-13-CDUP	9	-0.0328	-7470A	PPB	16 Nov 2020 15:11:48	U	SMPL
L4741-03MS - 1	11-13-CMS	14165	3.8454	-7470A	PPB	16 Nov 2020 15:14:03	U	SMPL
L4741-03MSD - 1	11-13-CMSD	14307	3.8843	-7470A	PPB	16 Nov 2020 15:16:19	U	SMPL
PB133039BL - 1	PBW	-54	-0.0501	-7470A	PPB	16 Nov 2020 15:18:35	U	SMPL
PB133039BS - 1	LCSW	13557	3.6788	-7470A	PPB	16 Nov 2020 15:20:50	U	SMPL
L4737-01 - 1	M-1-OW	398	0.0737	-7470A	PPB	16 Nov 2020 15:23:07	U	SMPL
L4737-01DUP - 1	M-1-OWDUP	258	0.0354	-7470A	PPB	16 Nov 2020 15:25:23	U	SMPL
L4737-01MS - 1	M-1-OWMS	15676	4.2593	-7470A	PPB	16 Nov 2020 15:27:40	U	SMPL
L4737-01MSD - 1	M-1-OWMSD	15870	4.3125	-7470A	PPB	16 Nov 2020 15:29:57	U	SMPL
L4737-02 - 1	MS-MSD	121	-0.0022	-7470A	PPB	16 Nov 2020 15:32:12	U	SMPL
CCV42 - 1	CCV42	18567	5.0513	-7470A	PPB	16 Nov 2020 15:34:28	U	SMPL
CCB42 - 1	CCB42	-100	-0.0627	-7470A	PPB	16 Nov 2020 15:36:43	U	SMPL
L4737-03 - 1	M-1-OW	98	-0.0085	-7470A	PPB	16 Nov 2020 15:38:59	U	SMPL
L4737-04 - 1	MS-MSD	41	-0.0241	-7470A	PPB	16 Nov 2020 15:41:14	U	SMPL
PB133038TB - 1	PB132979TB	-89	-0.0597	-7470A	PPB	16 Nov 2020 15:43:30	U	SMPL
L4741-03LX5 - 1		-23	-0.0416	-7470A	PPB	16 Nov 2020 15:45:47	U	SMPL
L4741-03A - 1		14628	3.9722	-7470A	PPB	16 Nov 2020 15:48:04	U	SMPL
L4737-01LX5 - 1		-45	-0.0476	-7470A	PPB	16 Nov 2020 15:50:20	U	SMPL
L4737-01A - 1		15759	4.2821	-7470A	PPB	16 Nov 2020 15:52:36	U	SMPL
CCV43 - 1	CCV43	18612	5.0637	-7470A	PPB	16 Nov 2020 15:54:53	U	SMPL
CCB43 - 1	CCB43	-80	-0.0572	-7470A	PPB	16 Nov 2020 15:57:10	U	SMPL