



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112000

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	11/16/2020	16:15
2	CAL2	1	Water	NA	NA	21.3	7.01	11/16/2020	16:16
3	CAL3	1	Water	NA	NA	21.3	10.03	11/16/2020	16:17
4	ICV	1	Water	NA	NA	21.3	6.99	11/16/2020	16:19
5	CCV1	1	Water	NA	NA	21.3	2.01	11/16/2020	16:20
6	L4766-04	1	Water	NA	NA	17.7	7.02	11/16/2020	16:21
7	L4766-04DUP	1	Water	NA	NA	17.8	7.03	11/16/2020	16:22
8	L4766-05	1	Water	NA	NA	17.6	6.91	11/16/2020	16:24
9	L4766-06	1	Water	NA	NA	17.8	6.94	11/16/2020	16:25
10	L4766-07	1	Water	NA	NA	17.5	6.90	11/16/2020	16:26
11	L4766-08	1	Water	NA	NA	17.6	7.07	11/16/2020	16:29
12	L4766-09	1	Water	NA	NA	18.2	6.99	11/16/2020	16:30
13	L4766-10	1	Water	NA	NA	18.3	7.61	11/16/2020	16:32
14	L4766-11	1	Water	NA	NA	18.3	6.10	11/16/2020	16:35
15	L4766-12	1	Water	NA	NA	18.2	7.00	11/16/2020	16:36
16	CCV2	1	Water	NA	NA	20.1	12.02	11/16/2020	16:38
17	L4767-02	1	Water	NA	NA	19.2	7.77	11/16/2020	16:40
18	L4769-11	1	Water	NA	NA	19.6	7.39	11/16/2020	16:41
19	CCV3	1	Water	NA	NA	20.2	2.01	11/16/2020	16:43

WORKLIST(Hardcopy Internal Chain)

20112000

WorkList Name : PH L4767 WorkList ID : 144195 Department : Wet-Chemistry Date : 11-16-2020 15:41:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2020	Water	L4766-04	pH	Cool 4 deg C	PSEG03	H11	31322	11/16/2020	9040C
11/16/2020	Water	L4766-05	pH	Cool 4 deg C	PSEG03	H11	31323	11/16/2020	9040C
11/16/2020	Water	L4766-06	pH	Cool 4 deg C	PSEG03	H11	31324	11/16/2020	9040C
11/16/2020	Water	L4766-07	pH	Cool 4 deg C	PSEG03	H11	31325	11/16/2020	9040C
11/16/2020	Water	L4766-08	pH	Cool 4 deg C	PSEG03	H11	31326	11/16/2020	9040C
11/16/2020	Water	L4766-09	pH	Cool 4 deg C	PSEG03	H11	31327	11/16/2020	9040C
11/16/2020	Water	L4766-10	pH	Cool 4 deg C	PSEG03	H11	31331	11/16/2020	9040C
11/16/2020	Water	L4766-11	pH	Cool 4 deg C	PSEG03	H11	31332	11/16/2020	9040C
11/16/2020	Water	L4766-12	pH	Cool 4 deg C	PSEG03	H11	TOTE-COMP-AUD	11/16/2020	9040C
11/16/2020	Water	L4767-02	pH	Cool 4 deg C	PSEG03	H11	NWB-1555	11/16/2020	9040C
11/16/2020	Water	L4769-11	pH	Cool 4 deg C	PSEG03	H11	TRE-1119	11/16/2020	9040C

Date/Time 11-16-2020 15:45
 Received by: JD (wet)
 Relinquished by: CP (SM)

Date/Time 11-16-2020 17:40
 Received by: CP (SM)
 Relinquished by: JD (wet)