



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112002

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	11/17/2020	10:05
2	CAL2	1	Water	NA	NA	21.3	7.02	11/17/2020	10:06
3	CAL3	1	Water	NA	NA	21.3	10.03	11/17/2020	10:07
4	ICV	1	Water	NA	NA	21.3	7.00	11/17/2020	10:08
5	CCV1	1	Water	NA	NA	21.2	2.01	11/17/2020	10:10
6	L4769-01	1	Solid	20.03	20	21.6	7.49	11/17/2020	10:11
7	L4769-01DUP	1	Solid	20.04	20	21.7	7.50	11/17/2020	10:12
8	L4769-04	1	Solid	20.02	20	21.4	7.59	11/17/2020	10:15
9	L4769-05	1	Solid	20.03	20	21.2	6.72	11/17/2020	10:16
10	L4769-06	1	Solid	20.01	20	21.3	7.11	11/17/2020	10:19
11	L4769-07	1	Solid	20.03	20	23.0	7.09	11/17/2020	10:20
12	L4769-10	1	Solid	20.03	20	20.9	11.85	11/17/2020	10:22
13	CCV2	1	Water	NA	NA	21.4	12.02	11/17/2020	10:24

WORKLIST(Hardcopy Internal Chain)

JB 712002

WorkList Name : ph 14769 solids WorkList ID : 144226 Department : Wet-Chemistry Date : 11-17-2020 09:59:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2020	Solid	L4769-01	pH	Cool 4 deg C	PSEG03	H11	TRE-1120-1122-COMP	11/16/2020	9045D
11/16/2020	Solid	L4769-04	pH	Cool 4 deg C	PSEG03	H11	1120-GRAB	11/16/2020	9045D
11/16/2020	Solid	L4769-05	pH	Cool 4 deg C	PSEG03	H11	1121-GRAB	11/16/2020	9045D
11/16/2020	Solid	L4769-06	pH	Cool 4 deg C	PSEG03	H11	1122-GRAB	11/16/2020	9045D
11/16/2020	Solid	L4769-07	pH	Cool 4 deg C	PSEG03	H11	TRE-1089	11/16/2020	9045D
11/16/2020	Solid	L4769-10	pH	Cool 4 deg C	PSEG03	H11	ETGI-359	11/16/2020	9045D

Date/Time 11-17-2020 10:00
 Received by: Jg (wetchem)
 Relinquished by: CP/SMI

Date/Time 11-17-2020 11:35
 Received by: CP SMI
 Relinquished by: Jg (wetchem)