



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112004

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2629         |
| buffer solution pH 7 yellow    | W2627         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 21.1             | 4.01        | 11/17/2020 | 12:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 21.1             | 7.00        | 11/17/2020 | 12:02     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 21.1             | 10.03       | 11/17/2020 | 12:03     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 21.1             | 6.98        | 11/17/2020 | 12:05     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 21.2             | 2.02        | 11/17/2020 | 12:06     |
| 6   | L4777-01    | 1  | Solid  | 20.03       | 20          | 21.8             | 8.68        | 11/17/2020 | 12:20     |
| 7   | L4777-01DUP | 1  | Solid  | 20.04       | 20          | 21.9             | 8.69        | 11/17/2020 | 12:21     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 21.2             | 12.02       | 11/17/2020 | 12:23     |

112004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4777      WorkList ID : 144229      Department : Wet-Chemistry      Date : 11-17-2020 11:48:39

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 11/17/2020 | Solid  | L4777-01 | pH   | Cool 4 deg C | PSEG03   | H12              | RT877           | 11/17/2020   | 9045D  |

Date/Time 11-17-2020 11:50  
Received by: JP Wet Chem  
Relinquished by: CP SMT

Date/Time 11-17-2020 13:00  
Received by: CP SMT  
Relinquished by: JP Wet Chem