



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112012

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	11/17/2020	16:50
2	CAL2	1	Water	NA	NA	21.3	7.01	11/17/2020	16:51
3	CAL3	1	Water	NA	NA	21.3	10.03	11/17/2020	16:52
4	ICV	1	Water	NA	NA	21.3	7.00	11/17/2020	17:00
5	CCV1	1	Water	NA	NA	21.3	2.02	11/17/2020	17:02
6	L4765-01	1	Solid	20.03	20	21.8	5.61	11/17/2020	17:05
7	L4765-03	1	Solid	20.02	20	21.6	5.54	11/17/2020	17:06
8	L4776-02	1	Solid	20.03	20	21.5	7.86	11/17/2020	17:10
9	L4777-01	1	Solid	20.02	20	21.8	8.68	11/17/2020	17:11
10	L4777-01DUP	1	Solid	20.03	20	21.9	8.69	11/17/2020	17:12
11	CCV2	1	Water	NA	NA	21.3	12.02	11/17/2020	17:15

WORKLIST(Hardcopy Internal Chain)

2012

WorkList Name : corrosivity I4755

WorkList ID : 144236

Department : Wet-Chemistry

Date : 11-17-2020 16:31:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2020	Solid	L4765-01	Corrosivity	Cool 4 deg C	PSEG03	H11	CF-OD	11/16/2020	9045D
11/16/2020	Solid	L4765-03	Corrosivity	Cool 4 deg C	PSEG03	H11	CF-OIL	11/16/2020	9045D
11/17/2020	Solid	L4776-02	Corrosivity	Cool 4 deg C	PSEG03	H12	ND-SOIL	11/17/2020	9045D
11/17/2020	Solid	L4777-01	Corrosivity	Cool 4 deg C	PSEG03	H12	RT877	11/17/2020	9045D

Date/Time 11/17/2020 16:35
 Received by: JA Laveren
 Relinquished by: JP

Date/Time 11/17/2020 18:00
 Received by: JP
 Relinquished by: JA Laveren