


LB112017

7470A

INSTRUMENT ID: LV1

μ Abs.:

36382



Linear

A= 0.0000e+000

B= 2.7566e-004

C= 8.3734e-005

Rho= 0.9999629

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	50	0.000	50				
0.2	0.200	0.197	-0.003	713	0.0 %	713				
2.5	2.500	2.482	-0.018	9002	0.0 %	9002				
5.0	5.000	5.037	0.037	18271	0.0 %	18271				
7.5	7.500	7.442	-0.058	26998	0.0 %	26998				
10.0	10.000	10.029	0.029	36382	0.0 %	36382				

LB112017 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Nov 2020 09:51:08				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	50	-	0.00007470A	PPB	18 Nov 2020 10:21:50	S	Std
0.2 - 1	50.2	713	-	0.20007470A	PPB	18 Nov 2020 10:24:07	S	Std
2.5 - 1	52.5	9002	-	2.50007470A	PPB	18 Nov 2020 10:26:23	S	Std
5.0 - 1	55	18271	-	5.00007470A	PPB	18 Nov 2020 10:28:40	S	Std
7.5 - 1	57.5	26998	-	7.50007470A	PPB	18 Nov 2020 10:30:57	S	Std
10.0 - 1	510	36382	-	10.00007470A	PPB	18 Nov 2020 10:33:12	S	Std
ICV08 - 1	ICV08	14426	3.9767	-7470A	PPB	18 Nov 2020 10:35:53	U	SMPL
ICB08 - 1	ICB08	-60	-0.0165	-7470A	PPB	18 Nov 2020 10:38:08	U	SMPL
CCV48 - 1	CCV48	18709	5.1574	-7470A	PPB	18 Nov 2020 10:45:23	U	SMPL
CCB48 - 1	CCB48	-69	-0.0189	-7470A	PPB	18 Nov 2020 10:47:38	U	SMPL
CRA - 1	CRA	702	0.1936	-7470A	PPB	18 Nov 2020 10:49:53	U	SMPL
HighStd - 1	HighStd	36995	10.1980	-7470A	PPB	18 Nov 2020 10:52:08	U	SMPL
ChkStd - 1	ChkStd	26218	7.2273	-7470A	PPB	18 Nov 2020 10:56:55	U	SMPL
PB133067BL - 1	PBW	-134	-0.0369	-7470A	PPB	18 Nov 2020 10:59:10	U	SMPL
PB133067BS - 1	LCSW	15257	4.2058	-7470A	PPB	18 Nov 2020 11:01:27	U	SMPL
L4767-03 - 1	NWB-1555	28	0.0078	-7470A	PPB	18 Nov 2020 11:03:43	U	SMPL
L4767-03DUP - 1	NWB-1555DUP	117	0.0323	-7470A	PPB	18 Nov 2020 11:06:00	U	SMPL
L4767-03MS - 1	NWB-1555MS	14403	3.9704	-7470A	PPB	18 Nov 2020 11:11:16	U	SMPL
L4767-03MSD - 1	NWB-1555MSD	14579	4.0189	-7470A	PPB	18 Nov 2020 11:13:33	U	SMPL
L4769-12 - 1	TRE-1119	-54	-0.0148	-7470A	PPB	18 Nov 2020 11:15:48	U	SMPL
CCV49 - 1	CCV49	18320	5.0501	-7470A	PPB	18 Nov 2020 11:18:03	U	SMPL
CCB49 - 1	CCB49	-45	-0.0123	-7470A	PPB	18 Nov 2020 11:20:18	U	SMPL
PB133068BL - 1	PBW	3	0.0009	-7470A	PPB	18 Nov 2020 11:22:33	U	SMPL
PB133068BS - 1	LCSW	14087	3.8833	-7470A	PPB	18 Nov 2020 11:24:49	U	SMPL
L4769-02 - 1	TRE-1120-1122-COMP	-32	-0.0087	-7470A	PPB	18 Nov 2020 11:27:04	U	SMPL
L4769-02DUP - 1	TRE-1120-1122-COMPDUP	52	0.0144	-7470A	PPB	18 Nov 2020 11:29:19	U	SMPL
L4769-02MS - 1	TRE-1120-1122-COMPMS	12787	3.5249	-7470A	PPB	18 Nov 2020 11:31:35	U	SMPL
L4769-02MSD - 1	TRE-1120-1122-COMPMSD	12741	3.5122	-7470A	PPB	18 Nov 2020 11:33:51	U	SMPL
L4769-08 - 1	TRE-1089	-14	-0.0038	-7470A	PPB	18 Nov 2020 11:36:08	U	SMPL
CCV50 - 1	CCV50	18489	5.0967	-7470A	PPB	18 Nov 2020 11:38:24	U	SMPL
CCB50 - 1	CCB50	-54	-0.0148	-7470A	PPB	18 Nov 2020 11:40:41	U	SMPL
PB133067TB - 1	PB133042TB	-3	-0.0007	-7470A	PPB	18 Nov 2020 11:42:57	U	SMPL
PB133068TB - 1	PB133060TB	8	0.0023	-7470A	PPB	18 Nov 2020 11:45:12	U	SMPL
L4767-03LX5 - 1		59	0.0163	-7470A	PPB	18 Nov 2020 11:47:27	U	SMPL
L4767-03A - 1		14557	4.0128	-7470A	PPB	18 Nov 2020 11:49:43	U	SMPL
L4769-02LX5 - 1		-34	-0.0093	-7470A	PPB	18 Nov 2020 11:51:58	U	SMPL
L4769-02A - 1		13478	3.7154	-7470A	PPB	18 Nov 2020 11:54:14	U	SMPL
CCV51 - 1	CCV51	18446	5.0849	-7470A	PPB	18 Nov 2020 11:56:30	U	SMPL
CCB51 - 1	CCB51	-49	-0.0134	-7470A	PPB	18 Nov 2020 11:58:45	U	SMPL