



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB112019

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/17/2020	16:45
2	CAL2	1	Water	NA	NA	20.1	7.01	11/17/2020	16:46
3	CAL3	1	Water	NA	NA	20.1	10.03	11/17/2020	16:47
4	ICV	1	Water	NA	NA	20.2	7.00	11/17/2020	16:48
5	CCV1	1	Water	NA	NA	20.2	2.02	11/17/2020	16:50
6	L4775-01	1	Water	NA	NA	13.2	7.66	11/17/2020	17:00
7	L4775-01DUP	1	Water	NA	NA	13.3	7.67	11/17/2020	17:01
8	L4778-08	1	Water	NA	NA	20.7	8.04	11/17/2020	17:02
9	L4778-11	1	Water	NA	NA	20.5	6.31	11/17/2020	17:05
10	L4779-01	1	Water	NA	NA	20.4	7.03	11/17/2020	17:06
11	L4779-02	1	Water	NA	NA	20.8	8.59	11/17/2020	17:09
12	L4779-03	1	Water	NA	NA	20.4	7.83	11/17/2020	17:10
13	L4779-04	1	Water	NA	NA	22.2	6.54	11/17/2020	17:11
14	L4779-05	1	Water	NA	NA	21.9	7.17	11/17/2020	17:12
15	L4779-06	1	Water	NA	NA	21.4	6.75	11/17/2020	17:15
16	CCV2	1	Water	NA	NA	21.1	12.02	11/17/2020	17:16
17	L4779-07	1	Water	NA	NA	21.0	6.68	11/17/2020	17:19
18	L4779-08	1	Water	NA	NA	21.4	6.69	11/17/2020	17:20
19	L4779-09	1	Water	NA	NA	21.4	6.67	11/17/2020	17:22
20	L4779-10	1	Water	NA	NA	20.7	7.37	11/17/2020	17:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L4779-11	1	Water	NA	NA	20.7	7.31	11/17/2020	17:26
22	L4779-12	1	Water	NA	NA	20.9	7.23	11/17/2020	17:29
23	L4779-13	1	Water	NA	NA	20.9	6.65	11/17/2020	17:30
24	L4779-14	1	Water	NA	NA	20.9	7.21	11/17/2020	17:32
25	L4779-15	1	Water	NA	NA	21.3	6.69	11/17/2020	18:00
26	CCV3	1	Water	NA	NA	21.2	2.02	11/17/2020	18:02

WORKLIST(Hardcopy Internal Chain)

JB 112019

WorkList Name : PH L4775

WorkList ID : 144235

Department : Wet-Chemistry

Date : 11-17-2020 16:30:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2020	Water	L4775-01	pH	Cool 4 deg C	PSEG03	H12	FMI-124	11/17/2020	9040C
11/17/2020	Water	L4778-08	pH	Cool 4 deg C	PSEG03	H12	1841	11/17/2020	9040C
11/17/2020	Water	L4778-11	pH	Cool 4 deg C	PSEG03	H12	1814	11/17/2020	9040C
11/17/2020	Water	L4779-01	pH	Cool 4 deg C	PSEG03	H12	1643	11/17/2020	9040C
11/17/2020	Water	L4779-02	pH	Cool 4 deg C	PSEG03	H12	1641	11/17/2020	9040C
11/17/2020	Water	L4779-03	pH	Cool 4 deg C	PSEG03	H12	1623	11/17/2020	9040C
11/17/2020	Water	L4779-04	pH	Cool 4 deg C	PSEG03	H12	1608	11/17/2020	9040C
11/17/2020	Water	L4779-05	pH	Cool 4 deg C	PSEG03	H12	1555	11/17/2020	9040C
11/17/2020	Water	L4779-06	pH	Cool 4 deg C	PSEG03	H12	1654	11/17/2020	9040C
11/17/2020	Water	L4779-07	pH	Cool 4 deg C	PSEG03	H12	1655	11/17/2020	9040C
11/17/2020	Water	L4779-08	pH	Cool 4 deg C	PSEG03	H12	1656	11/17/2020	9040C
11/17/2020	Water	L4779-09	pH	Cool 4 deg C	PSEG03	H12	1657	11/17/2020	9040C
11/17/2020	Water	L4779-10	pH	Cool 4 deg C	PSEG03	H12	1605	11/17/2020	9040C
11/17/2020	Water	L4779-11	pH	Cool 4 deg C	PSEG03	H12	1606	11/17/2020	9040C
11/17/2020	Water	L4779-12	pH	Cool 4 deg C	PSEG03	H12	1607	11/17/2020	9040C
11/17/2020	Water	L4779-13	pH	Cool 4 deg C	PSEG03	H12	COMP-1654-1658	11/17/2020	9040C
11/17/2020	Water	L4779-14	pH	Cool 4 deg C	PSEG03	H12	COMP-1605-1607	11/17/2020	9040C
11/17/2020	Water	L4779-15	pH	Cool 4 deg C	PSEG03	H12	1658	11/17/2020	9040C

Date/Time 11-17-2020 16:35
 Received by: JB (vet Chem)
 Relinquished by: JB (SM)

Date/Time 11-17-2020 18:35
 Received by: JB (SM)
 Relinquished by: JB (vet Chem)