



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112020

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/17/2020	16:45
2	CAL2	1	Water	NA	NA	20.1	7.01	11/17/2020	16:46
3	CAL3	1	Water	NA	NA	20.1	10.03	11/17/2020	16:47
4	ICV	1	Water	NA	NA	20.2	7.00	11/17/2020	16:48
5	CCV1	1	Water	NA	NA	20.2	2.02	11/17/2020	16:50
6	L4775-01	1	Water	NA	NA	13.2	7.66	11/17/2020	17:00
7	L4775-01DUP	1	Water	NA	NA	13.3	7.67	11/17/2020	17:01
8	CCV2	1	Water	NA	NA	21.1	12.02	11/17/2020	17:16

WORKLIST(Hardcopy Internal Chain)

JB 112020

WorkList Name : corrosivity I4776 WorkList ID : 144237 Department : Wet-Chemistry Date : 11-17-2020 16:31:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2020	Water	L4775-01	Corrosivity	Cool 4 deg C	PSEG03	H12	FMI-124	11/17/2020	9040C

Date/Time 11-17-2020 16:35
 Received by: JB (wet-chem)
 Relinquished by: JB (S.M.)

Date/Time 11-17-2020 18:35
 Received by: JB (S.M.)
 Relinquished by: JB (wet-chem)